

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032519\
 Data File : VX008417.D
 Acq On : 25 Mar 2019 17:47
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX032519

Quant Time: Mar 26 08:03:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 04:58:03 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	118	0.00
2 T	Dichlorodifluoromethane	0.589	0.585	0.7	128	0.00
3 P	Chloromethane	0.841	0.780	7.3	123	0.00
4 C	Vinyl Chloride	0.823	0.800	2.8#	123	0.00
5 T	Bromomethane	0.492	0.431	12.4	125	0.00
6 T	Chloroethane	0.487	0.491	-0.8	121	0.00
7 T	Trichlorofluoromethane	0.932	0.925	0.8	125	0.00
8 T	Diethyl Ether	0.441	0.430	2.5	122	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.563	0.583	-3.6	130	0.00
10 T	Methyl Iodide	0.505	0.566	-12.1	127	0.00
11 T	Tert butyl alcohol	0.168	0.156	7.1	110	0.00
12 CM	1,1-Dichloroethene	0.582	0.574	1.4#	123	0.00
13 T	Acrolein	0.122	0.163	-33.6#	170#	0.00
14 T	Allyl chloride	1.282	1.263	1.5	122	0.00
15 T	Acrylonitrile	0.440	0.396	10.0	112	0.00
16 T	Acetone	0.396	0.368	7.1	115	0.00
17 T	Carbon Disulfide	1.930	1.731	10.3	124	0.00
18 T	Methyl Acetate	0.952	0.848	10.9	113	0.00
19 T	Methyl tert-butyl Ether	2.096	2.101	-0.2	122	0.00
20 T	Methylene Chloride	0.713	0.678	4.9	123	0.00
21 T	trans-1,2-Dichloroethene	0.637	0.609	4.4	124	0.00
22 T	Diisopropyl ether	2.523	2.494	1.1	122	0.00
23 T	Vinyl Acetate	2.202	2.205	-0.1	120	0.00
24 P	1,1-Dichloroethane	1.274	1.261	1.0	123	0.00
25 T	2-Butanone	0.635	0.588	7.4	113	0.00
26 T	2,2-Dichloropropane	0.847	0.907	-7.1	129	0.00
27 T	cis-1,2-Dichloroethene	0.711	0.705	0.8	123	0.00
28 T	Bromochloromethane	0.629	0.589	6.4	119	0.00
29 T	Tetrahydrofuran	0.419	0.371	11.5	111	0.00
30 C	Chloroform	1.150	1.144	0.5#	123	0.00
31 T	Cyclohexane	1.215	1.244	-2.4	128	0.00
32 T	1,1,1-Trichloroethane	0.919	0.942	-2.5	124	0.00
33 S	1,2-Dichloroethane-d4	0.758	0.734	3.2	117	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	114	0.00
35 S	Dibromofluoromethane	0.326	0.335	-2.8	118	0.00
36 T	1,1-Dichloropropene	0.533	0.558	-4.7	126	0.00
37 T	Ethyl Acetate	0.691	0.678	1.9	114	0.00
38 T	Carbon Tetrachloride	0.445	0.482	-8.3	125	0.00
39 T	Methylcyclohexane	0.637	0.714	-12.1	132	0.00
40 TM	Benzene	1.596	1.660	-4.0	123	0.00
41 T	Methacrylonitrile	0.388	0.385	0.8	115	0.00
42 TM	1,2-Dichloroethane	0.584	0.587	-0.5	122	0.00
43 T	Isopropyl Acetate	1.105	1.116	-1.0	117	0.00
44 TM	Trichloroethene	0.375	0.386	-2.9	125	0.00
45 C	1,2-Dichloropropane	0.454	0.473	-4.2#	122	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032519\
 Data File : VX008417.D
 Acq On : 25 Mar 2019 17:47
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX032519

Quant Time: Mar 26 08:03:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 04:58:03 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.265	0.271	-2.3	123	0.00
47 T	Bromodichloromethane	0.507	0.551	-8.7	125	0.00
48 T	Methyl methacrylate	0.553	0.577	-4.3	119	0.00
49 T	1,4-Dioxane	0.011	0.011	0.0	111	0.00
50 S	Toluene-d8	1.304	1.353	-3.8	118	0.00
51 T	4-Methyl-2-Pentanone	0.717	0.727	-1.4	115	0.00
52 CM	Toluene	0.954	1.014	-6.3#	125	0.00
53 T	t-1,3-Dichloropropene	0.555	0.633	-14.1	127	0.00
54 T	cis-1,3-Dichloropropene	0.632	0.703	-11.2	126	0.00
55 T	1,1,2-Trichloroethane	0.388	0.401	-3.4	121	0.00
56 T	Ethyl methacrylate	0.701	0.755	-7.7	122	0.00
57 T	1,3-Dichloropropane	0.696	0.718	-3.2	122	0.00
58 T	2-Chloroethyl Vinyl ether	0.362	0.299	17.4	93	0.00
59 T	2-Hexanone	0.556	0.568	-2.2	114	0.00
60 T	Dibromochloromethane	0.369	0.411	-11.4	125	0.00
61 T	1,2-Dibromoethane	0.396	0.415	-4.8	121	0.00
62 S	4-Bromofluorobenzene	0.483	0.523	-8.3	123	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	116	0.00
64 T	Tetrachloroethene	0.338	0.354	-4.7	128	0.00
65 PM	Chlorobenzene	1.087	1.141	-5.0	126	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.403	-7.5	125	0.00
67 C	Ethyl Benzene	2.052	2.170	-5.8#	126	0.00
68 T	m/p-Xylenes	0.747	0.791	-5.9	126	0.00
69 T	o-Xylene	0.724	0.766	-5.8	125	0.00
70 T	Styrene	1.251	1.373	-9.8	128	0.00
71 P	Bromoform	0.296	0.326	-10.1	124	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
73 T	Isopropylbenzene	4.068	4.159	-2.2	129	0.00
74 T	N-amyl acetate	2.354	2.380	-1.1	119	0.00
75 P	1,1,2,2-Tetrachloroethane	1.578	1.479	6.3	118	0.00
76 T	1,2,3-Trichloropropane	1.486	1.261	15.1	103	0.00
77 T	Bromobenzene	0.987	0.982	0.5	127	0.00
78 T	n-propylbenzene	4.765	5.011	-5.2	131	0.00
79 T	2-Chlorotoluene	2.898	2.927	-1.0	127	0.00
80 T	1,3,5-Trimethylbenzene	3.439	3.578	-4.0	130	0.00
81 T	trans-1,4-Dichloro-2-butene	0.464	0.500	-7.8	122	0.00
82 T	4-Chlorotoluene	3.343	3.459	-3.5	131	0.00
83 T	tert-Butylbenzene	3.318	3.458	-4.2	131	0.00
84 T	1,2,4-Trimethylbenzene	3.478	3.592	-3.3	129	0.00
85 T	sec-Butylbenzene	3.951	4.196	-6.2	132	0.00
86 T	p-Isopropyltoluene	3.486	3.793	-8.8	133	0.00
87 T	1,3-Dichlorobenzene	1.764	1.803	-2.2	131	0.00
88 T	1,4-Dichlorobenzene	1.780	1.817	-2.1	131	0.00
89 T	n-Butylbenzene	3.303	3.730	-12.9	137	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032519\
Data File : VX008417.D
Acq On : 25 Mar 2019 17:47
Operator : JC/SP
Sample : VSTDICV050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX032519

Quant Time: Mar 26 08:03:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 04:58:03 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.581	0.642	-10.5	134	0.00
91 T	1,2-Dichlorobenzene	1.766	1.783	-1.0	127	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.344	0.317	7.8	114	0.00
93 T	1,2,4-Trichlorobenzene	1.189	1.276	-7.3	133	0.00
94 T	Hexachlorobutadiene	0.565	0.623	-10.3	135	0.00
95 T	Naphthalene	4.077	4.029	1.2	121	0.00
96 T	1,2,3-Trichlorobenzene	1.195	1.244	-4.1	130	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6