

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021501.D
 Acq On : 25 Mar 2021 20:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 26 03:19:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	79493	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	132192	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	123576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	58459	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	64078	48.45	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.90%
35) Dibromofluoromethane	5.464	113	45467	48.42	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.84%
50) Toluene-d8	8.694	98	163871	48.97	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.94%
62) 4-Bromofluorobenzene	11.118	95	62435	47.99	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.98%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.187	85	43501	49.79	ug/l	100
3) Chloromethane	1.311	50	52703	52.47	ug/l	97
4) Vinyl Chloride	1.393	62	46665	52.22	ug/l	99
5) Bromomethane	1.623	94	19425	54.01	ug/l	97
6) Chloroethane	1.699	64	28735	50.70	ug/l	99
7) Trichlorofluoromethane	1.911	101	74080	50.97	ug/l	99
8) Diethyl Ether	2.170	74	25404	50.37	ug/l	76
9) 1,1,2-Trichlorotrifluo...	2.364	101	40979	50.08	ug/l	96
10) Methyl Iodide	2.487	142	34490	53.45	ug/l #	88
11) Tert butyl alcohol	3.017	59	49678	227.19	ug/l	97
12) 1,1-Dichloroethene	2.352	96	39373	49.15	ug/l	86
13) Acrolein	2.276	56	25237	229.18	ug/l	100
14) Allyl chloride	2.705	41	80442	49.47	ug/l #	89
15) Acrylonitrile	3.117	53	136667	262.99	ug/l	98
16) Acetone	2.423	43	128246	256.44	ug/l	88
17) Carbon Disulfide	2.546	76	102514	45.57	ug/l	99
18) Methyl Acetate	2.746	43	80943	56.02	ug/l	90
19) Methyl tert-butyl Ether	3.164	73	156803	52.41	ug/l	97
20) Methylene Chloride	2.835	84	47569	48.74	ug/l #	87
21) trans-1,2-Dichloroethene	3.140	96	44592	50.68	ug/l	91
22) Diisopropyl ether	3.829	45	165454	52.37	ug/l	92
23) Vinyl Acetate	3.788	43	701085	262.36	ug/l #	92
24) 1,1-Dichloroethane	3.670	63	86824	50.98	ug/l	97
25) 2-Butanone	4.635	43	198873	262.33	ug/l #	89
26) 2,2-Dichloropropane	4.547	77	62402	42.29	ug/l	97
27) cis-1,2-Dichloroethene	4.558	96	53026	51.89	ug/l	94
28) Bromochloromethane	4.982	49	41795	49.02	ug/l	85
29) Tetrahydrofuran	5.088	42	127933	260.28	ug/l	86
30) Chloroform	5.176	83	94371	52.50	ug/l	99
31) Cyclohexane	5.541	56	72178	49.63	ug/l	93
32) 1,1,1-Trichloroethane	5.464	97	82419	51.74	ug/l	96
36) 1,1-Dichloropropene	5.764	75	64998	50.18	ug/l	97
37) Ethyl Acetate	4.794	43	74681	49.86	ug/l #	93
38) Carbon Tetrachloride	5.753	117	70687	50.79	ug/l	99
39) Methylcyclohexane	7.435	83	66884	46.58	ug/l	91
40) Benzene	6.111	78	190398	51.77	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021501.D
 Acq On : 25 Mar 2021 20:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 26 03:19:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	5.000	41	45369	52.23	ug/l #	88
42) 1,2-Dichloroethane	6.159	62	82631	51.82	ug/l	94
43) Isopropyl Acetate	6.411	43	126458	52.00	ug/l #	91
44) Trichloroethene	7.188	130	48941	50.42	ug/l	96
45) 1,2-Dichloropropane	7.488	63	51842	52.03	ug/l	96
46) Dibromomethane	7.635	93	36829	50.71	ug/l	94
47) Bromodichloromethane	7.876	83	75330	50.97	ug/l	97
48) Methyl methacrylate	7.741	41	67436	51.98	ug/l #	83
49) 1,4-Dioxane	7.712	88	18336	789.52	ug/l #	88
51) 4-Methyl-2-Pentanone	8.618	43	405363	271.01	ug/l	88
52) Toluene	8.765	92	118632	52.22	ug/l	97
53) t-1,3-Dichloropropene	9.024	75	75906	50.86	ug/l	96
54) cis-1,3-Dichloropropene	8.412	75	80824	50.89	ug/l #	84
55) 1,1,2-Trichloroethane	9.194	97	50625	52.80	ug/l	97
56) Ethyl methacrylate	9.159	69	78194	48.35	ug/l #	76
57) 1,3-Dichloropropane	9.353	76	85844	52.32	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.288	63	209943	259.08	ug/l	95
59) 2-Hexanone	9.471	43	302080	268.96	ug/l	85
60) Dibromochloromethane	9.565	129	56796	52.89	ug/l	99
61) 1,2-Dibromoethane	9.653	107	53707	52.14	ug/l	98
64) Tetrachloroethene	9.318	164	42935	51.87	ug/l	93
65) Chlorobenzene	10.118	112	126810	51.08	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.206	131	49653	52.20	ug/l	100
67) Ethyl Benzene	10.236	91	233246	51.55	ug/l	100
68) m/p-Xylenes	10.341	106	166907	103.53	ug/l	92
69) o-Xylene	10.683	106	81100	51.94	ug/l	93
70) Styrene	10.694	104	140749	52.41	ug/l	95
71) Bromoform	10.841	173	39519	51.12	ug/l #	100
73) Isopropylbenzene	11.000	105	217486	52.95	ug/l	100
74) N-amyl acetate	10.883	43	108635	54.66	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	11.253	83	77205	53.20	ug/l	100
76) 1,2,3-Trichloropropane	11.277	75	87428	64.24	ug/l	91
77) Bromobenzene	11.236	156	54359	53.39	ug/l	94
78) n-propylbenzene	11.342	91	254864	52.54	ug/l	98
79) 2-Chlorotoluene	11.406	91	153531	52.18	ug/l	99
80) 1,3,5-Trimethylbenzene	11.489	105	189257	53.93	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.059	75	22144	49.56	ug/l #	83
82) 4-Chlorotoluene	11.495	91	182310	52.18	ug/l	97
83) tert-Butylbenzene	11.753	119	172799	50.58	ug/l	91
84) 1,2,4-Trimethylbenzene	11.789	105	190932	53.97	ug/l	97
85) sec-Butylbenzene	11.930	105	199017	50.50	ug/l	100
86) p-Isopropyltoluene	12.047	119	185973	51.84	ug/l	97
87) 1,3-Dichlorobenzene	12.006	146	94906	51.76	ug/l	98
88) 1,4-Dichlorobenzene	12.077	146	94532	49.84	ug/l	96
89) n-Butylbenzene	12.371	91	161383	49.10	ug/l	98
90) Hexachloroethane	12.577	117	32356	49.98	ug/l	92
91) 1,2-Dichlorobenzene	12.371	146	90401	50.71	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.983	75	18172	49.70	ug/l	84
93) 1,2,4-Trichlorobenzene	13.624	180	58149	49.77	ug/l	97
94) Hexachlorobutadiene	13.765	225	22923	43.47	ug/l	96
95) Naphthalene	13.818	128	205349	53.88	ug/l	99
96) 1,2,3-Trichlorobenzene	14.001	180	58167	49.80	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021501.D
Acq On : 25 Mar 2021 20:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Mar 26 03:19:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 26 03:19:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.623 min Scan# 771
Delta R.T. -0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 79493
Ion Ratio Lower Upper
168 100
99 63.1 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 49.79 ug/l
RT: 1.187 min Scan# 17
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 85 Resp: 43501
Ion Ratio Lower Upper
85 100
87 32.6 16.4 49.2









#3
Chloromethane
Concen: 52.47 ug/l
RT: 1.311 min Scan# 38
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 52703
Ion Ratio Lower Upper
50 100
52 31.6 26.4 39.6



#4
Vinyl Chloride
Concen: 52.22 ug/l
RT: 1.393 min Scan# 52
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 62 Resp: 46665
Ion Ratio Lower Upper
62 100
64 31.6 25.7 38.5





#5

Bromomethane

Concen: 54.01 ug/l

RT: 1.623 min Scan# 91

Delta R.T. 0.001 min

Lab File: VX021501.D





Acq: 25 Mar 2021 20:59

Tgt Ion: 94 Resp: 19425

Ion Ratio Lower Upper

94 100

96 98.1 76.4 114.6



#6

Chloroethane

Concen: 50.70 ug/l

RT: 1.699 min Scan# 104

Delta R.T. -0.006 min

Lab File: VX021501.D




Acq: 25 Mar 2021 20:59


Tgt Ion: 64 Resp: 28735

Ion Ratio Lower Upper

64 100

66 32.2 25.4 38.0









#7
Trichlorofluoromethane
Concen: 50.97 ug/l
RT: 1.911 min Scan# 140
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 74080
Ion Ratio Lower Upper
101 100
103 65.6 53.2 79.8



#8
Diethyl Ether
Concen: 50.37 ug/l
RT: 2.170 min Scan# 184
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 74 Resp: 25404
Ion Ratio Lower Upper
74 100
45 111.7 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.08 ug/l
RT: 2.364 min Scan# 217
Delta R.T. 0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 40979
Ion Ratio Lower Upper
101 100
85 47.6 34.9 52.3
151 74.2 61.4 92.2



#10
Methyl Iodide
Concen: 53.45 ug/l
RT: 2.487 min Scan# 238
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:142 Resp: 34490
Ion Ratio Lower Upper
142 100
127 51.6 33.3 49.9#
141 14.5 11.4 17.0









#11
Tert butyl alcohol
Concen: 227.19 ug/l
RT: 3.017 min Scan# 328
Delta R.T. -0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 59 Resp: 49678
Ion Ratio Lower Upper
59 100
57 10.8 7.8 11.8



#12
1,1-Dichloroethene
Concen: 49.15 ug/l
RT: 2.352 min Scan# 215
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 96 Resp: 39373
Ion Ratio Lower Upper
96 100
61 178.5 124.6 187.0
98 66.3 50.6 76.0









#13
Acrolein
Concen: 229.18 ug/l
RT: 2.276 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 25237
Ion Ratio Lower Upper
56 100
55 70.9 56.4 84.6



#14
Allyl chloride
Concen: 49.47 ug/l
RT: 2.705 min Scan# 275
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 41 Resp: 80442
Ion Ratio Lower Upper
41 100
39 71.4 51.6 77.4
76 28.8 30.5 45.7#









#15
Acrylonitrile
Concen: 262.99 ug/l
RT: 3.117 min Scan# 345
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 53 Resp: 136667
Ion Ratio Lower Upper
53 100
52 83.1 65.3 97.9
51 37.3 28.2 42.2



#16
Acetone
Concen: 256.44 ug/l
RT: 2.423 min Scan# 227
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 43 Resp: 128246
Ion Ratio Lower Upper
43 100
58 28.0 27.8 41.8









#17
Carbon Disulfide
Concen: 45.57 ug/l
RT: 2.546 min Scan# 248
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 102514
Ion Ratio Lower Upper
76 100
78 8.8 7.2 10.8



#18
Methyl Acetate
Concen: 56.02 ug/l
RT: 2.746 min Scan# 282
Delta R.T. -0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 43 Resp: 80943
Ion Ratio Lower Upper
43 100
74 22.1 21.9 32.9





#19

Methyl tert-butyl Ether

Concen: 52.41 ug/l

RT: 3.164 min Scan# 353

Delta R.T. 0.000 min

Lab File: VX021501.D





Acq: 25 Mar 2021 20:59

Tgt Ion: 73 Resp: 156803

Ion Ratio Lower Upper

73 100

57 22.7 17.1 25.7



#20

Methylene Chloride

Concen: 48.74 ug/l

RT: 2.835 min Scan# 297

Delta R.T. 0.001 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Tgt Ion: 84 Resp: 47569




Ion Ratio Lower Upper


84 100

49 131.9 88.6 132.8

51 41.7 27.0 40.6#


86 63.5 50.3 75.5








#21
trans-1,2-Dichloroethene
Concen: 50.68 ug/l
RT: 3.140 min Scan# 349
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 44592
Ion Ratio Lower Upper
96 100
61 148.5 107.7 161.5
98 64.8 50.4 75.6




#22
Diisopropyl ether
Concen: 52.37 ug/l
RT: 3.829 min Scan# 466
Delta R.T. 0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59



Tgt Ion: 45 Resp: 165454
Ion Ratio Lower Upper
45 100
43 65.1 58.0 87.0
87 26.5 25.4 38.0
59 11.0 10.2 15.2









#23
 Vinyl Acetate
 Concen: 262.36 ug/l
 RT: 3.788 min Scan# 459
 Delta R.T. 0.006 min
 Lab File: VX021501.D
 Acq: 25 Mar 2021 20:59




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 701085
 Ion Ratio Lower Upper
 43 100
 86 9.3 9.9 14.9#



#24
 1,1-Dichloroethane
 Concen: 50.98 ug/l
 RT: 3.670 min Scan# 439
 Delta R.T. -0.000 min
 Lab File: VX021501.D
 Acq: 25 Mar 2021 20:59

Tgt Ion: 63 Resp: 86824
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.8 11.2
 100 3.6 2.5 7.5









#25
2-Butanone
Concen: 262.33 ug/l
RT: 4.635 min Scan# 603
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 198873
Ion Ratio Lower Upper
43 100
72 23.6 23.7 35.5#



#26
2,2-Dichloropropane
Concen: 42.29 ug/l
RT: 4.547 min Scan# 588
Delta R.T. -0.005 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 77 Resp: 62402
Ion Ratio Lower Upper
77 100
97 21.9 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 51.89 ug/l

RT: 4.558 min Scan# 590





Delta R.T. 0.000 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Tgt Ion: 96 Resp: 53026

Ion	Ratio	Lower	Upper
96	100		
61	152.7	0.0	285.6
98	63.9	0.0	129.6



#28

Bromochloromethane

Concen: 49.02 ug/l

RT: 4.982 min Scan# 662




Delta R.T. 0.006 min


Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Tgt Ion: 49 Resp: 41795

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.0
130	71.2	68.2	102.2





#29

Tetrahydrofuran

Concen: 260.28 ug/l

RT: 5.088 min Scan# 680

Delta R.T. -0.000 min





Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 42 Resp: 127933

Ion	Ratio	Lower	Upper
42	100		
72	42.1	41.7	62.5
71	39.2	38.4	57.6



#30

Chloroform

Concen: 52.50 ug/l

RT: 5.176 min Scan# 695




Delta R.T. 0.000 min


Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Tgt Ion: 83 Resp: 94371

Ion	Ratio	Lower	Upper
83	100		
85	64.1	52.0	78.0









#31
Cyclohexane
Concen: 49.63 ug/l
RT: 5.541 min Scan# 757
Delta R.T. -0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 72178
Ion Ratio Lower Upper
56 100
69 29.8 25.3 37.9
84 81.3 71.4 107.0



#32
1,1,1-Trichloroethane
Concen: 51.74 ug/l
RT: 5.464 min Scan# 744
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 97 Resp: 82419
Ion Ratio Lower Upper
97 100
99 64.6 51.5 77.3
61 48.3 34.2 51.4





#33

1,2-Dichloroethane-d4

Concen: 48.45 ug/l

RT: 6.029 min Scan# 840

Delta R.T. 0.000 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Tgt Ion: 65 Resp: 64078

Ion Ratio Lower Upper





65 100

67 50.0 0.0 105.4

Instrument :

MSVOA_X

ClientSampled :



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.829 min Scan# 976

Delta R.T. 0.000 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59




Tgt Ion: 114 Resp: 132192


Ion Ratio Lower Upper

114 100

63 23.5 0.0 40.8

88 16.6 0.0 32.8





#35

Dibromofluoromethane

Concen: 48.42 ug/l

RT: 5.464 min Scan# 744

Delta R.T. 0.000 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Tgt Ion:113 Resp: 45467

Ion Ratio Lower Upper

113 100





111 104.0 81.9 122.9

192 19.2 16.7 25.1

Instrument :

MSVOA_X

ClientSampleId :



#36

1,1-Dichloropropene

Concen: 50.18 ug/l

RT: 5.764 min Scan# 795

Delta R.T. 0.000 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59




Tgt Ion: 75 Resp: 64998


Ion Ratio Lower Upper

75 100

110 33.7 18.1 54.4

77 30.4 25.3 37.9





#37

Ethyl Acetate

Concen: 49.86 ug/l

RT: 4.794 min Scan# 630





Delta R.T. 0.001 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Tgt Ion:	43	Resp:	74681
Ion	Ratio	Lower	Upper
43	100		
61	13.4	12.8	19.2
70	10.2	10.5	15.7

Instrument :
MSVOA_X
ClientSampleId :



#38

Carbon Tetrachloride

Concen: 50.79 ug/l




RT: 5.753 min Scan# 793


Delta R.T. 0.001 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Tgt Ion:	117	Resp:	70687
Ion	Ratio	Lower	Upper
117	100		
119	96.2	76.2	114.2
121	29.5	24.8	37.2





#39

Methylcyclohexane

Concen: 46.58 ug/l

RT: 7.435 min Scan# 1079

Delta R.T. 0.000 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59





Tgt Ion: 83 Resp: 66884

Ion Ratio Lower Upper

83 100

55 84.0 58.9 88.3

98 44.9 38.1 57.1



#40

Benzene

Concen: 51.77 ug/l

RT: 6.111 min Scan# 854

Delta R.T. 0.000 min

Lab File: VX021501.D




Acq: 25 Mar 2021 20:59


Tgt Ion: 78 Resp: 190398

Ion Ratio Lower Upper

78 100

77 23.7 18.9 28.3









#41
Methacrylonitrile
Concen: 52.23 ug/l
RT: 5.000 min Scan# 665
Delta R.T. -0.005 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	41	Resp:	45369
Ion	Ratio	Lower	Upper
41	100		
39	58.1	42.5	63.7
67	64.3	65.0	97.6#
52	29.9	25.3	37.9



#42
1,2-Dichloroethane
Concen: 51.82 ug/l
RT: 6.159 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:	62	Resp:	82631
Ion	Ratio	Lower	Upper
62	100		
98	7.9	0.0	20.2









#43
Isopropyl Acetate
Concen: 52.00 ug/l
RT: 6.411 min Scan# 905
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 126458
Ion Ratio Lower Upper
43 100
61 18.1 18.2 27.4#
87 11.9 12.5 18.7#



#44
Trichloroethene
Concen: 50.42 ug/l
RT: 7.188 min Scan# 1037
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:130 Resp: 48941
Ion Ratio Lower Upper
130 100
95 103.4 0.0 198.0





#45

1,2-Dichloropropane

Concen: 52.03 ug/l

RT: 7.488 min Scan# 1088

Delta R.T. 0.000 min

Lab File: VX021501.D





Acq: 25 Mar 2021 20:59

Tgt Ion: 63 Resp: 51842

Ion Ratio Lower Upper

63 100

65 29.8 25.7 38.5



#46

Dibromomethane

Concen: 50.71 ug/l

RT: 7.635 min Scan# 1113

Delta R.T. 0.000 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59




Tgt Ion: 93 Resp: 36829


Ion Ratio Lower Upper

93 100

95 84.1 66.8 100.2

174 92.8 83.1 124.7









#47
Bromodichloromethane
Concen: 50.97 ug/l
RT: 7.876 min Scan# 1154
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 75330
Ion Ratio Lower Upper
83 100
85 62.0 51.4 77.0
127 8.8 6.8 10.2



#48
Methyl methacrylate
Concen: 51.98 ug/l
RT: 7.741 min Scan# 1131
Delta R.T. -0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 41 Resp: 67436
Ion Ratio Lower Upper
41 100
69 74.9 76.0 114.0#
39 63.8 44.6 67.0









#49
1,4-Dioxane
Concen: 789.52 ug/l
RT: 7.712 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 88 Resp: 18336
Ion Ratio Lower Upper
88 100
43 38.7 23.2 34.8#
58 75.8 55.0 82.4



#50
Toluene-d8
Concen: 48.97 ug/l
RT: 8.694 min Scan# 1293
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 98 Resp: 163871
Ion Ratio Lower Upper
98 100
100 65.2 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 271.01 ug/l
RT: 8.618 min Scan# 1280
Delta R.T. 0.001 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 405363
Ion Ratio Lower Upper
43 100
58 36.6 35.4 53.0



#52
Toluene
Concen: 52.22 ug/l
RT: 8.765 min Scan# 1305
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 92 Resp: 118632
Ion Ratio Lower Upper
92 100
91 175.3 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 50.86 ug/l
RT: 9.024 min Scan# 1349
Delta R.T. 0.001 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 75906
Ion Ratio Lower Upper
75 100
77 33.3 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 50.89 ug/l
RT: 8.412 min Scan# 1245
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59


Tgt Ion: 75 Resp: 80824
Ion Ratio Lower Upper
75 100
77 31.8 25.7 38.5
39 61.8 35.1 52.7#








#55
1,1,2-Trichloroethane
Concen: 52.80 ug/l
RT: 9.194 min Scan# 1378
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion	Ratio	Lower	Upper
97	100		
83	89.9	69.3	103.9
85	57.2	44.3	66.5
99	61.2	50.3	75.5



#56
Ethyl methacrylate
Concen: 48.35 ug/l
RT: 9.159 min Scan# 1372
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion	Ratio	Lower	Upper
69	100		
41	74.7	46.8	70.2#
39	46.3	24.7	37.1#









#57
1,3-Dichloropropane
Concen: 52.32 ug/l
RT: 9.353 min Scan# 1405
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 85844
Ion Ratio Lower Upper
76 100
78 32.4 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 259.08 ug/l
RT: 8.288 min Scan# 1224
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 63 Resp: 209943
Ion Ratio Lower Upper
63 100
106 26.8 23.5 35.3









#59
2-Hexanone
Concen: 268.96 ug/l
RT: 9.471 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 302080
Ion Ratio Lower Upper
43 100
58 50.5 30.9 92.8



#60
Dibromochloromethane
Concen: 52.89 ug/l
RT: 9.565 min Scan# 1441
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 129 Resp: 56796
Ion Ratio Lower Upper
129 100
127 77.8 38.4 115.2









#61
1,2-Dibromoethane
Concen: 52.14 ug/l
RT: 9.653 min Scan# 1456
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 53707
Ion Ratio Lower Upper
107 100
109 92.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 47.99 ug/l
RT: 11.118 min Scan# 1705
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 95 Resp: 62435
Ion Ratio Lower Upper
95 100
174 73.7 0.0 154.6
176 71.6 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.094 min Scan# 1531
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 123576
Ion Ratio Lower Upper
117 100
82 59.1 47.4 71.2
119 32.7 26.6 39.8



#64
Tetrachloroethene
Concen: 51.87 ug/l
RT: 9.318 min Scan# 1399
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:164 Resp: 42935
Ion Ratio Lower Upper
164 100
166 125.0 103.4 155.2
129 102.8 70.8 106.2
131 94.7 72.5 108.7









#65
Chlorobenzene
Concen: 51.08 ug/l
RT: 10.118 min Scan# 1535
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:112 Resp: 126810
Ion Ratio Lower Upper
112 100
114 31.3 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 52.20 ug/l
RT: 10.206 min Scan# 1550
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:131 Resp: 49653
Ion Ratio Lower Upper
131 100
133 95.1 47.7 143.1
119 63.9 32.1 96.5









#67
Ethyl Benzene
Concen: 51.55 ug/l
RT: 10.236 min Scan# 1555
Delta R.T. 0.001 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 233246
Ion Ratio Lower Upper
91 100
106 31.0 24.6 37.0



#68
m/p-Xylenes
Concen: 103.53 ug/l
RT: 10.341 min Scan# 1573
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:106 Resp: 166907
Ion Ratio Lower Upper
106 100
91 216.1 162.8 244.2









#69
o-Xylene
Concen: 51.94 ug/l
RT: 10.683 min Scan# 1631
Delta R.T. 0.001 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 81100
Ion Ratio Lower Upper
106 100
91 226.5 108.1 324.2



#70
Styrene
Concen: 52.41 ug/l
RT: 10.694 min Scan# 1633
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:104 Resp: 140749
Ion Ratio Lower Upper
104 100
78 55.6 40.0 60.0
103 55.3 43.4 65.0









#71
Bromoform
Concen: 51.12 ug/l
RT: 10.841 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 39519
Ion Ratio Lower Upper
173 100
175 48.2 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.059 min Scan# 1865
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:152 Resp: 58459
Ion Ratio Lower Upper
152 100
115 88.0 41.1 123.3
150 175.6 0.0 349.0





#73

Isopropylbenzene

Concen: 52.95 ug/l

RT: 11.000 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VX021501.D





Acq: 25 Mar 2021 20:59

Tgt Ion:105 Resp: 217486

Ion Ratio Lower Upper

105 100

120 26.6 13.2 39.6



#74

N-ethyl acetate

Concen: 54.66 ug/l

RT: 10.883 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX021501.D

Acq: 25 Mar 2021 20:59

Tgt Ion: 43 Resp: 108635




Ion Ratio Lower Upper


43 100

70 38.1 39.0 58.6#

55 23.4 21.5 32.3

61 21.4 22.1 33.1#









#75
1,1,2,2-Tetrachloroethane
Concen: 53.20 ug/l
RT: 11.253 min Scan# 1728
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 77205
Ion Ratio Lower Upper
83 100
131 9.7 5.1 15.2
85 63.5 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 64.24 ug/l
RT: 11.277 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 75 Resp: 87428
Ion Ratio Lower Upper
75 100
77 32.7 19.1 57.5









#77
Bromobenzene
Concen: 53.39 ug/l
RT: 11.236 min Scan# 1725
Delta R.T. 0.001 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 54359
Ion Ratio Lower Upper
156 100
77 174.1 82.8 248.6
158 93.6 49.1 147.4



#78
n-propylbenzene
Concen: 52.54 ug/l
RT: 11.342 min Scan# 1743
Delta R.T. 0.001 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 91 Resp: 254864
Ion Ratio Lower Upper
91 100
120 22.1 11.5 34.5









#79
2-Chlorotoluene
Concen: 52.18 ug/l
RT: 11.406 min Scan# 1754
Delta R.T. 0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 153531
Ion Ratio Lower Upper
91 100
126 32.7 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 53.93 ug/l
RT: 11.489 min Scan# 1768
Delta R.T. 0.001 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 105 Resp: 189257
Ion Ratio Lower Upper
105 100
120 46.4 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 49.56 ug/l
RT: 11.059 min Scan# 1695
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 22144
Ion Ratio Lower Upper
75 100
53 117.5 74.5 111.7#
89 48.5 37.9 56.9



#82
4-Chlorotoluene
Concen: 52.18 ug/l
RT: 11.495 min Scan# 1769
Delta R.T. 0.001 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 91 Resp: 182310
Ion Ratio Lower Upper
91 100
126 28.3 14.9 44.5









#83
tert-Butylbenzene
Concen: 50.58 ug/l
RT: 11.753 min Scan# 1813
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:119 Resp: 172799
Ion Ratio Lower Upper
119 100
91 64.5 27.9 83.7
134 22.7 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 53.97 ug/l
RT: 11.789 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:105 Resp: 190932
Ion Ratio Lower Upper
105 100
120 42.7 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 50.50 ug/l
 RT: 11.930 min Scan# 1843
 Delta R.T. -0.000 min
 Lab File: VX021501.D
 Acq: 25 Mar 2021 20:59




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion:105 Resp: 199017
 Ion Ratio Lower Upper
 105 100
 134 19.8 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 51.84 ug/l
 RT: 12.047 min Scan# 1863
 Delta R.T. 0.000 min
 Lab File: VX021501.D
 Acq: 25 Mar 2021 20:59

Tgt Ion:119 Resp: 185973
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.9 38.7
 91 26.6 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 51.76 ug/l
RT: 12.006 min Scan# 1856
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 94906
Ion Ratio Lower Upper
146 100
111 42.5 20.2 60.5
148 62.8 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 49.84 ug/l
RT: 12.077 min Scan# 1868
Delta R.T. -0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:146 Resp: 94532
Ion Ratio Lower Upper
146 100
111 42.0 19.6 58.8
148 65.5 31.6 95.0









#89
n-Butylbenzene
Concen: 49.10 ug/l
RT: 12.371 min Scan# 1918
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 161383
Ion Ratio Lower Upper
91 100
92 53.0 27.4 82.0
134 24.9 13.2 39.5



#90
Hexachloroethane
Concen: 49.98 ug/l
RT: 12.577 min Scan# 1953
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 117 Resp: 32356
Ion Ratio Lower Upper
117 100
201 73.9 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 50.71 ug/l
RT: 12.371 min Scan# 1918
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 90401
Ion Ratio Lower Upper
146 100
111 45.2 20.8 62.2
148 64.2 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.70 ug/l
RT: 12.983 min Scan# 2022
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion: 75 Resp: 18172
Ion Ratio Lower Upper
75 100
155 75.1 45.3 135.9
157 98.8 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 49.77 ug/l
RT: 13.624 min Scan# 2131
Delta R.T. -0.006 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:180 Resp: 58149
Ion Ratio Lower Upper
180 100
182 98.3 47.7 143.1
145 32.5 15.0 45.0



#94
Hexachlorobutadiene
Concen: 43.47 ug/l
RT: 13.765 min Scan# 2155
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:225 Resp: 22923
Ion Ratio Lower Upper
225 100
223 63.8 31.1 93.2
227 67.2 31.6 95.0









#95
Naphthalene
Concen: 53.88 ug/l
RT: 13.818 min Scan# 2164
Delta R.T. 0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:128 Resp: 205349
Ion Ratio Lower Upper
128 100
127 13.3 10.3 15.5
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 49.80 ug/l
RT: 14.001 min Scan# 2195
Delta R.T. -0.000 min
Lab File: VX021501.D
Acq: 25 Mar 2021 20:59

Tgt Ion:180 Resp: 58167
Ion Ratio Lower Upper
180 100
182 95.5 47.6 142.9
145 35.1 16.1 48.3

