

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021484.D
 Acq On : 25 Mar 2021 14:38
 Operator : JC/MD
 Sample : M1724-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P1679

Quant Time: Mar 25 15:54:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	58775	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	107895	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109902	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	49692	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	51618	52.78	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.56%	
35) Dibromofluoromethane	5.464	113	38454	50.17	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.34%	
50) Toluene-d8	8.700	98	139816	51.19	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.38%	
62) 4-Bromofluorobenzene	11.123	95	56215	52.94	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.88%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.316	50	4041	5.48	ug/l	97
11) Tert butyl alcohol	3.016	59	10049	62.16	ug/l	96
16) Acetone	2.422	43	10772	29.13	ug/l	90
17) Carbon Disulfide	2.552	76	5176	3.11	ug/l	96
19) Methyl tert-butyl Ether	3.175	73	1168	0.53	ug/l	95
20) Methylene Chloride	2.834	84	980	1.36	ug/l #	78
25) 2-Butanone	4.640	43	2960	5.28	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021484.D
Acq On : 25 Mar 2021 14:38
Operator : JC/MD
Sample : M1724-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P1679

Quant Time: Mar 25 15:54:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

