

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021486.D
 Acq On : 25 Mar 2021 15:19
 Operator : JC/MD
 Sample : M1758-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 26 03:14:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	63903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	115406	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	118279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52041	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	53992	50.78	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.56%
35) Dibromofluoromethane	5.464	113	40929	49.93	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.86%
50) Toluene-d8	8.694	98	147242	50.40	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.80%
62) 4-Bromofluorobenzene	11.118	95	59924	52.76	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.52%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021486.D
Acq On : 25 Mar 2021 15:19
Operator : JC/MD
Sample : M1758-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 26 03:14:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

