

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021487.D
 Acq On : 25 Mar 2021 15:40
 Operator : JC/MD
 Sample : M1758-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 26 03:15:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	63299	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	116389	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	117351	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	50800	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	54537	51.78	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.56%
35) Dibromofluoromethane	5.464	113	41203	49.84	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.68%
50) Toluene-d8	8.694	98	146706	49.80	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.60%
62) 4-Bromofluorobenzene	11.118	95	58056	50.69	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.38%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021487.D
Acq On : 25 Mar 2021 15:40
Operator : JC/MD
Sample : M1758-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 26 03:15:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

