

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021490.D
 Acq On : 25 Mar 2021 16:43
 Operator : JC/MD
 Sample : M1787-02DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D2-20210318DL

Quant Time: Mar 26 03:15:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	60483	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	107098	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	113898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	53015	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	51352	51.03	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.06%		
35) Dibromofluoromethane	5.464	113	38653	50.81	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.62%		
50) Toluene-d8	8.694	98	140298	51.75	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 103.50%		
62) 4-Bromofluorobenzene	11.118	95	58931	55.91	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 111.82%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	4494	7.22	ug/l	95
12) 1,1-Dichloroethene	2.358	96	2511	4.12	ug/l	98
27) cis-1,2-Dichloroethene	4.564	96	571	0.73	ug/l	81
32) 1,1,1-Trichloroethane	5.458	97	570	0.47	ug/l #	45
38) Carbon Tetrachloride	5.752	117	420	0.37	ug/l #	94
44) Trichloroethene	7.194	130	51285	65.22	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021490.D
Acq On : 25 Mar 2021 16:43
Operator : JC/MD
Sample : M1787-02DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D2-20210318DL

Quant Time: Mar 26 03:15:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

