

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021491.D
 Acq On : 25 Mar 2021 17:05
 Operator : JC/MD
 Sample : M1787-06DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-103D2-20210318DL

Quant Time: Mar 26 03:16:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	60605	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	110257	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	113899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	53101	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51319	50.89	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.78%	
35) Dibromofluoromethane	5.464	113	39335	50.22	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.44%	
50) Toluene-d8	8.694	98	141706	50.77	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.54%	
62) 4-Bromofluorobenzene	11.118	95	59895	55.20	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.40%	
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.358	101	411	0.66	ug/l	Qvalue # 80
44) Trichloroethene	7.188	130	49808	61.53	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021491.D
Acq On : 25 Mar 2021 17:05
Operator : JC/MD
Sample : M1787-06DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-103D2-20210318DL

Quant Time: Mar 26 03:16:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

