

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021492.D
 Acq On : 25 Mar 2021 17:29
 Operator : JC/MD
 Sample : M1787-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-103D1-20210318DL

Quant Time: Mar 26 03:16:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	57336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	106219	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109652	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48150	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	50021	52.43	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.86%
35) Dibromofluoromethane	5.464	113	37729	50.01	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.02%
50) Toluene-d8	8.694	98	135787	50.50	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.00%
62) 4-Bromofluorobenzene	11.118	95	53846	51.51	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.02%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.358	101	417	0.71	ug/l #	84
12) 1,1-Dichloroethene	2.364	96	200	0.35	ug/l #	68
44) Trichloroethene	7.188	130	44233	56.72	ug/l	95
64) Tetrachloroethene	9.318	164	475	0.65	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021492.D
Acq On : 25 Mar 2021 17:29
Operator : JC/MD
Sample : M1787-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-103D1-20210318DL

Quant Time: Mar 26 03:16:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

