

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021493.D
 Acq On : 25 Mar 2021 17:52
 Operator : JC/MD
 Sample : M1787-08DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-125D1-20210319DL

Quant Time: Mar 26 03:16:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	62126	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	111831	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52923	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	53162	51.43	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.86%	
35) Dibromofluoromethane	5.464	113	39950	50.29	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.58%	
50) Toluene-d8	8.694	98	143920	50.84	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.68%	
62) 4-Bromofluorobenzene	11.118	95	59149	53.75	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.50%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.358	101	1831	2.86	ug/l	94
12) 1,1-Dichloroethene	2.346	96	391	0.62	ug/l #	77
27) cis-1,2-Dichloroethene	4.564	96	883	1.11	ug/l	88
44) Trichloroethene	7.188	130	52474	63.91	ug/l	97
64) Tetrachloroethene	9.318	164	1641	2.11	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021493.D
Acq On : 25 Mar 2021 17:52
Operator : JC/MD
Sample : M1787-08DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-125D1-20210319DL

Quant Time: Mar 26 03:16:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

