

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021495.D
 Acq On : 25 Mar 2021 18:39
 Operator : JC/MD
 Sample : M1787-10DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-125D3-20210319DL

Quant Time: Mar 26 03:17:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	59257	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.829	114	110752	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110758	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48569	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52580	53.33	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.66%	
35) Dibromofluoromethane	5.458	113	38789	49.31	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.62%	
50) Toluene-d8	8.694	98	141483	50.47	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.94%	
62) 4-Bromofluorobenzene	11.118	95	55727	51.13	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.26%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	13300	21.81	ug/l	95
27) cis-1,2-Dichloroethene	4.558	96	469	0.62	ug/l	83
44) Trichloroethene	7.188	130	37965	46.69	ug/l	100
64) Tetrachloroethene	9.318	164	1231	1.66	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021495.D
Acq On : 25 Mar 2021 18:39
Operator : JC/MD
Sample : M1787-10DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-125D3-20210319DL

Quant Time: Mar 26 03:17:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

