

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021497.D
 Acq On : 25 Mar 2021 19:26
 Operator : JC/MD
 Sample : M1787-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-122D3-20210319

Quant Time: Mar 26 03:18:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	60619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	110475	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115677	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	53848	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52596	52.15	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.30%
35) Dibromofluoromethane	5.464	113	40109	51.11	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.22%
50) Toluene-d8	8.694	98	143758	51.41	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.82%
62) 4-Bromofluorobenzene	11.118	95	59109	54.37	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.74%
Target Compounds						
					Qvalue	
16) Acetone	2.417	43	1516	3.98	ug/l #	86
44) Trichloroethene	7.182	130	1786	2.20	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021497.D
Acq On : 25 Mar 2021 19:26
Operator : JC/MD
Sample : M1787-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-122D3-20210319

Quant Time: Mar 26 03:18:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

