

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021507.D
 Acq On : 26 Mar 2021 00:07
 Operator : JC/MD
 Sample : M1770-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-03242021

Quant Time: Mar 26 06:32:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	60861	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	112145	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115397	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	52896	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51953	51.30	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.60%
35) Dibromofluoromethane	5.464	113	39273	49.30	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.60%
50) Toluene-d8	8.694	98	141511	49.85	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.70%
62) 4-Bromofluorobenzene	11.118	95	58325	52.85	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.70%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021507.D
Acq On : 26 Mar 2021 00:07
Operator : JC/MD
Sample : M1770-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-03242021

Quant Time: Mar 26 06:32:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

