

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021508.D
 Acq On : 26 Mar 2021 00:30
 Operator : JC/MD
 Sample : M1817-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2021-03-24-TB-01

Quant Time: Mar 26 06:33:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	62033	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	113126	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109799	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	44236	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52783	51.14	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.28%
35) Dibromofluoromethane	5.464	113	40864	50.85	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.70%
50) Toluene-d8	8.694	98	140763	49.16	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.32%
62) 4-Bromofluorobenzene	11.118	95	53245	47.83	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.66%
Target Compounds					Qvalue	
16) Acetone	2.423	43	19730	50.56	ug/l	89
25) 2-Butanone	4.646	43	2711	4.58	ug/l #	84
37) Ethyl Acetate	4.799	43	1390	1.08	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021508.D
Acq On : 26 Mar 2021 00:30
Operator : JC/MD
Sample : M1817-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2021-03-24-TB-01

Quant Time: Mar 26 06:33:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

