

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021509.D
 Acq On : 26 Mar 2021 00:54
 Operator : JC/MD
 Sample : M1770-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-103S

Quant Time: Mar 26 06:33:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	63919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	117125	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	117596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52001	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	55919	52.58	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.16%			
35) Dibromofluoromethane	5.464	113	42679	51.30	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.60%			
50) Toluene-d8	8.694	98	145877	49.20	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 98.40%			
62) 4-Bromofluorobenzene	11.118	95	59081	51.26	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.52%			
Target Compounds						Qvalue
11) Tert butyl alcohol	3.017	59	1580	8.99	ug/l	99
16) Acetone	2.422	43	4546	11.31	ug/l #	86
40) Benzene	6.111	78	6624	2.03	ug/l	93
95) Naphthalene	13.818	128	24391	7.20	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021509.D
Acq On : 26 Mar 2021 00:54
Operator : JC/MD
Sample : M1770-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-103S

Quant Time: Mar 26 06:33:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

