

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021510.D
 Acq On : 26 Mar 2021 01:17
 Operator : JC/MD
 Sample : M1770-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-102D

Quant Time: Mar 26 06:33:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	61507	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	112804	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	118206	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	55308	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	53190	51.97	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.94%		
35) Dibromofluoromethane	5.464	113	40448	50.48	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.96%		
50) Toluene-d8	8.694	98	143662	50.31	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 100.62%		
62) 4-Bromofluorobenzene	11.118	95	60538	54.53	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 109.06%		
Target Compounds						Qvalue
16) Acetone	2.417	43	2172	5.61	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021510.D
Acq On : 26 Mar 2021 01:17
Operator : JC/MD
Sample : M1770-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-102D

Quant Time: Mar 26 06:33:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

