

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021514.D
 Acq On : 26 Mar 2021 02:51
 Operator : JC/MD
 Sample : M1806-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW4

Quant Time: Mar 26 06:34:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	59870	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109159	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109890	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	47068	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51908	52.11	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.22%	
35) Dibromofluoromethane	5.458	113	38992	50.29	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.58%	
50) Toluene-d8	8.694	98	139137	50.36	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.72%	
62) 4-Bromofluorobenzene	11.118	95	54064	50.33	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.66%	
Target Compounds						
16) Acetone	2.417	43	2216	5.88	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021514.D
Acq On : 26 Mar 2021 02:51
Operator : JC/MD
Sample : M1806-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW4

Quant Time: Mar 26 06:34:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

