

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021520.D
 Acq On : 26 Mar 2021 05:12
 Operator : JC/MD
 Sample : M1777-07 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 243

Quant Time: Mar 26 06:36:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	64328	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	122426	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	116543	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	50568	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	56064	52.38	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.76%
35) Dibromofluoromethane	5.464	113	42648	49.04	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.08%
50) Toluene-d8	8.694	98	150960	48.71	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.42%
62) 4-Bromofluorobenzene	11.118	95	58959	48.94	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.88%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.999	59	2583	14.60	ug/l #	84
16) Acetone	2.423	43	8144	20.12	ug/l #	87
68) m/p-Xylenes	10.341	106	1147	0.75	ug/l	90
84) 1,2,4-Trimethylbenzene	11.788	105	1242	0.41	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021520.D
Acq On : 26 Mar 2021 05:12
Operator : JC/MD
Sample : M1777-07 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
243

Quant Time: Mar 26 06:36:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

