

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021521.D
 Acq On : 26 Mar 2021 05:36
 Operator : JC/MD
 Sample : M1829-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7B

Quant Time: Mar 26 06:37:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	65814	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	124482	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	125337	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	58748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	57319	52.34	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.68%
35) Dibromofluoromethane	5.464	113	43509	49.21	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.42%
50) Toluene-d8	8.694	98	157286	49.92	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.84%
62) 4-Bromofluorobenzene	11.118	95	65275	53.28	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.56%
Target Compounds						
16) Acetone	2.423	43	1930	4.66	ug/l	87
44) Trichloroethene	7.188	130	298	0.33	ug/l	84
64) Tetrachloroethene	9.318	164	121280	144.45	ug/l	97
68) m/p-Xylenes	10.341	106	849	0.52	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021521.D
Acq On : 26 Mar 2021 05:36
Operator : JC/MD
Sample : M1829-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7B

Quant Time: Mar 26 06:37:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

