

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021523.D
 Acq On : 26 Mar 2021 06:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 26 06:43:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.550	0.544	1.1	89	0.00
3 P	Chloromethane	0.702	0.617	12.1	92	0.00
4 C	Vinyl Chloride	0.562	0.578	-2.8#	92	0.00
5 T	Bromomethane	0.226	0.230	-1.8	95	0.00
6 T	Chloroethane	0.357	0.365	-2.2	93	0.00
7 T	Trichlorofluoromethane	0.914	0.969	-6.0	94	0.00
8 T	Diethyl Ether	0.317	0.331	-4.4	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.515	0.535	-3.9	95	0.00
10 T	Methyl Iodide	0.362	0.330	8.8	83	0.00
11 T	Tert butyl alcohol	0.138	0.139	-0.7	95	0.00
12 CM	1,1-Dichloroethene	0.504	0.507	-0.6#	96	0.00
13 T	Acrolein	0.077	0.060	22.1	82	0.00
14 T	Allyl chloride	1.023	0.993	2.9	89	0.00
15 T	Acrylonitrile	0.327	0.350	-7.0	96	0.00
16 T	Acetone	0.315	0.336	-6.7	100	0.00
17 T	Carbon Disulfide	1.415	1.316	7.0	89	0.00
18 T	Methyl Acetate	0.909	1.050	-15.5	102	0.00
19 T	Methyl tert-butyl Ether	1.882	2.044	-8.6	98	0.00
20 T	Methylene Chloride	0.614	0.622	-1.3	95	0.00
21 T	trans-1,2-Dichloroethene	0.553	0.573	-3.6	95	0.00
22 T	Diisopropyl ether	1.987	2.121	-6.7	96	0.00
23 T	Vinyl Acetate	1.681	1.805	-7.4	95	0.00
24 P	1,1-Dichloroethane	1.071	1.131	-5.6	96	0.00
25 T	2-Butanone	0.477	0.512	-7.3	98	0.00
26 T	2,2-Dichloropropane	0.928	0.605	34.8#	58	0.00
27 T	cis-1,2-Dichloroethene	0.643	0.669	-4.0	95	0.00
28 T	Bromochloromethane	0.536	0.535	0.2	96	0.00
29 T	Tetrahydrofuran	0.309	0.331	-7.1	97	0.00
30 C	Chloroform	1.131	1.225	-8.3#	97	0.00
31 T	Cyclohexane	0.915	0.933	-2.0	91	0.00
32 T	1,1,1-Trichloroethane	1.002	1.063	-6.1	95	0.00
33 S	1,2-Dichloroethane-d4	0.832	0.840	-1.0	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.355	0.346	2.5	97	0.00
36 T	1,1-Dichloropropene	0.490	0.491	-0.2	94	0.00
37 T	Ethyl Acetate	0.566	0.574	-1.4	98	0.00
38 T	Carbon Tetrachloride	0.526	0.544	-3.4	94	0.00
39 T	Methylcyclohexane	0.543	0.524	3.5	87	0.00
40 TM	Benzene	1.391	1.440	-3.5	95	0.00
41 T	Methacrylonitrile	0.329	0.335	-1.8	97	0.00
42 TM	1,2-Dichloroethane	0.603	0.624	-3.5	96	0.00
43 T	Isopropyl Acetate	0.920	0.955	-3.8	96	0.00
44 TM	Trichloroethene	0.367	0.375	-2.2	96	0.00
45 C	1,2-Dichloropropane	0.377	0.395	-4.8#	96	0.00
46 T	Dibromomethane	0.275	0.283	-2.9	94	0.00
47 T	Bromodichloromethane	0.559	0.581	-3.9	95	0.00
48 T	Methyl methacrylate	0.491	0.508	-3.5	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
 Data File : VX021523.D
 Acq On : 26 Mar 2021 06:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 26 06:43:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	90	0.00
50 S	Toluene-d8	1.266	1.209	4.5	92	0.00
51 T	4-Methyl-2-Pentanone	0.566	0.608	-7.4	96	0.00
52 CM	Toluene	0.859	0.886	-3.1#	93	0.00
53 T	t-1,3-Dichloropropene	0.564	0.537	4.8	85	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.585	2.7	87	0.00
55 T	1,1,2-Trichloroethane	0.363	0.377	-3.9	94	0.00
56 T	Ethyl methacrylate	0.548	0.579	-5.7	92	0.00
57 T	1,3-Dichloropropane	0.621	0.642	-3.4	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.307	0.317	-3.3	92	0.00
59 T	2-Hexanone	0.425	0.458	-7.8	97	0.00
60 T	Dibromochloromethane	0.406	0.430	-5.9	94	0.00
61 T	1,2-Dibromoethane	0.390	0.406	-4.1	93	0.00
62 S	4-Bromofluorobenzene	0.492	0.470	4.5	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.335	0.356	-6.3	94	0.00
65 PM	Chlorobenzene	1.004	1.010	-0.6	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.405	-5.2	93	0.00
67 C	Ethyl Benzene	1.831	1.916	-4.6#	92	0.00
68 T	m/p-Xylenes	0.652	0.694	-6.4	92	0.00
69 T	o-Xylene	0.632	0.654	-3.5	91	0.00
70 T	Styrene	1.086	1.143	-5.2	92	0.00
71 P	Bromoform	0.313	0.335	-7.0	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.513	3.905	-11.2	93	0.00
74 T	N-amyl acetate	1.700	1.855	-9.1	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.241	1.371	-10.5	97	0.00
76 T	1,2,3-Trichloropropane	1.164	1.301	-11.8	93	0.00
77 T	Bromobenzene	0.871	0.917	-5.3	90	0.00
78 T	n-propylbenzene	4.149	4.464	-7.6	91	0.00
79 T	2-Chlorotoluene	2.517	2.748	-9.2	93	0.00
80 T	1,3,5-Trimethylbenzene	3.001	3.278	-9.2	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.382	0.346	9.4	78	0.00
82 T	4-Chlorotoluene	2.988	3.209	-7.4	91	0.00
83 T	tert-Butylbenzene	2.922	3.094	-5.9	90	0.00
84 T	1,2,4-Trimethylbenzene	3.026	3.353	-10.8	93	0.00
85 T	sec-Butylbenzene	3.371	3.623	-7.5	89	0.00
86 T	p-Isopropyltoluene	3.068	3.253	-6.0	86	0.00
87 T	1,3-Dichlorobenzene	1.568	1.688	-7.7	94	0.00
88 T	1,4-Dichlorobenzene	1.622	1.675	-3.3	93	0.00
89 T	n-Butylbenzene	2.811	2.884	-2.6	85	0.00
90 T	Hexachloroethane	0.554	0.597	-7.8	93	0.00
91 T	1,2-Dichlorobenzene	1.525	1.699	-11.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.286	0.333	-16.4	99	0.00
93 T	1,2,4-Trichlorobenzene	0.999	1.070	-7.1	93	0.00
94 T	Hexachlorobutadiene	0.451	0.414	8.2	81	0.00
95 T	Naphthalene	3.259	3.766	-15.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032521\
Data File : VX021523.D
Acq On : 26 Mar 2021 06:23
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 26 06:43:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.999	1.057	-5.8	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6