



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032522\
 Data File : VX027694.D
 Acq On : 25 Mar 2022 11:13
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Title : SW846 8260
 Last Update : Fri Mar 25 16:26:59 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	3296	PASS
75	95	30	60	57.2	9126	PASS
95	95	100	100	100.0	15962	PASS
96	95	5	9	6.6	1047	PASS
173	174	0.00	2	0.4	41	PASS
174	95	50	100	60.4	9639	PASS
175	174	5	9	7.6	731	PASS
176	174	95	101	97.4	9385	PASS
177	176	5	9	7.0	658	PASS