

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032524\
 Data File : VX040785.D
 Acq On : 25 Mar 2024 09:56
 Operator : JC/MD
 Sample : VX0325WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0325WBL01

Quant Time: Mar 26 04:44:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	48319	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	85196	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	77735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	38554	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	39854	49.053	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.100%
35) Dibromofluoromethane	5.385	113	29406	46.128	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.260%
50) Toluene-d8	8.647	98	104085	47.684	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.360%
62) 4-Bromofluorobenzene	11.079	95	40766	47.271	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032524\
Data File : VX040785.D
Acq On : 25 Mar 2024 09:56
Operator : JC/MD
Sample : VX0325WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0325WBL01

Quant Time: Mar 26 04:44:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration

