

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032524\
 Data File : VX040784.D
 Acq On : 25 Mar 2024 09:33
 Operator : JC/MD
 Sample : VX0325MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0325MBL01

Quant Time: Mar 26 04:43:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	45790	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	82776	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	68958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41018	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	38087	49.468	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.940%
35) Dibromofluoromethane	5.379	113	28341	45.757	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.520%
50) Toluene-d8	8.647	98	98684	46.532	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.060%
62) 4-Bromofluorobenzene	11.079	95	38668	46.149	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.300%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032524\
Data File : VX040784.D
Acq On : 25 Mar 2024 09:33
Operator : JC/MD
Sample : VX0325MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0325MBL01

Quant Time: Mar 26 04:43:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:44:48 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 712
Delta R.T. 0.006 min
Lab File: VX040784.D
Acq: 25 Mar 2024 09:33




Instrument :
MSVOA_X
ClientSampleId :
VX0325MBL01


Tgt Ion:168 Resp: 45790
Ion Ratio Lower Upper
168 100
99 61.4 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 49.468 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX040784.D
Acq: 25 Mar 2024 09:33

Tgt Ion: 65 Resp: 38087
Ion Ratio Lower Upper
65 100
67 50.1 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX040784.D
Acq: 25 Mar 2024 09:33




Instrument :
MSVOA_X
ClientSampleId :
VX0325MBL01


Tgt Ion:114 Resp: 82776
Ion Ratio Lower Upper
114 100
63 21.6 0.0 45.0
88 17.0 0.0 36.6



#35
Dibromofluoromethane
Concen: 45.757 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX040784.D
Acq: 25 Mar 2024 09:33

Tgt Ion:113 Resp: 28341
Ion Ratio Lower Upper
113 100
111 101.0 82.9 124.3
192 17.5 13.3 19.9









#50
Toluene-d8
Concen: 46.532 ug/l
RT: 8.647 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX040784.D
Acq: 25 Mar 2024 09:33




Instrument :
MSVOA_X
ClientSampleId :
VX0325MBL01


Tgt Ion: 98 Resp: 98684
Ion Ratio Lower Upper
98 100
100 63.6 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 46.149 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX040784.D
Acq: 25 Mar 2024 09:33

Tgt Ion: 95 Resp: 38668
Ion Ratio Lower Upper
95 100
174 75.6 0.0 135.6
176 72.0 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX040784.D
Acq: 25 Mar 2024 09:33

Instrument :
MSVOA_X
ClientSampleId :
VX0325MBL01

Tgt Ion:117 Resp: 68958
Ion Ratio Lower Upper
117 100
82 59.4 46.2 69.4
119 33.2 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.006 min
Lab File: VX040784.D
Acq: 25 Mar 2024 09:33

Tgt Ion:152 Resp: 41018
Ion Ratio Lower Upper
152 100
115 61.6 43.3 129.9
150 154.5 0.0 345.0

