



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032524\
Data File : VX040782.D
Acq On : 25 Mar 2024 08:28
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032124W.M
Title : SW846 8260
Last Update : Fri Mar 22 02:44:48 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.8	3948	PASS
75	95	30	60	57.5	10408	PASS
95	95	100	100	100.0	18102	PASS
96	95	5	9	6.0	1086	PASS
173	174	0.00	2	0.7	86	PASS
174	95	50	100	72.3	13081	PASS
175	174	5	9	7.3	958	PASS
176	174	95	101	95.2	12453	PASS
177	176	5	9	6.3	782	PASS