

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032525\  
 Data File : VX045452.D  
 Acq On : 25 Mar 2025 19:51  
 Operator : JC/MD  
 Sample : Q1608-11  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE132D1-20250318

Quant Time: Mar 27 03:17:47 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X022825W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 28 06:45:16 2025  
 Response via : Initial Calibration


| Compound                      | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|-------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                         |                |      |                     |         |        |          |
| Internal Standards            |                |      |                     |         |        |          |
| 1) Pentafluorobenzene         | 5.544          | 168  | 72182               | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene       | 6.757          | 114  | 140177              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5          | 10.049         | 117  | 126304              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4    | 12.018         | 152  | 48163               | 50.000  | ug/l   | 0.00     |
|                               |                |      |                     |         |        |          |
| System Monitoring Compounds   |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4     | 5.952          | 65   | 60650               | 52.824  | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 74 - 125 |      | Recovery = 105.640% |         |        |          |
| 35) Dibromofluoromethane      | 5.379          | 113  | 45356               | 48.389  | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 75 - 124 |      | Recovery = 96.780%  |         |        |          |
| 50) Toluene-d8                | 8.647          | 98   | 157998              | 46.495  | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 86 - 113 |      | Recovery = 93.000%  |         |        |          |
| 62) 4-Bromofluorobenzene      | 11.079         | 95   | 54366               | 48.281  | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 77 - 121 |      | Recovery = 96.560%  |         |        |          |
|                               |                |      |                     |         |        |          |
| Target Compounds              |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane    | 1.166          | 85   | 241                 | 0.265   | ug/l   | 78       |
| 5) Bromomethane               | 1.593          | 94   | 128                 | 0.295   | ug/l # | 3        |
| 9) 1,1,2-Trichlorotrifluo...  | 2.319          | 101  | 4131                | 4.924   | ug/l   | 99       |
| 11) Tert butyl alcohol        | 2.934          | 59   | 515                 | 2.368   | ug/l # | 73       |
| 12) 1,1-Dichloroethene        | 2.307          | 96   | 652                 | 0.735   | ug/l   | 91       |
| 16) Acetone                   | 2.386          | 43   | 926                 | 1.738   | ug/l   | 94       |
| 19) Methyl tert-butyl Ether   | 3.117          | 73   | 2254                | 0.757   | ug/l   | 99       |
| 22) Diisopropyl ether         | 3.764          | 45   | 1502                | 0.468   | ug/l # | 74       |
| 23) Vinyl Acetate             | 3.757          | 43   | 997                 | 0.372   | ug/l # | 73       |
| 24) 1,1-Dichloroethane        | 3.605          | 63   | 932                 | 0.518   | ug/l # | 83       |
| 27) cis-1,2-Dichloroethene    | 4.501          | 96   | 2326                | 2.150   | ug/l   | 88       |
| 28) Bromochloromethane        | 5.093          | 49   | 241                 | 0.281   | ug/l # | 17       |
| 30) Chloroform                | 5.093          | 83   | 3811                | 2.131   | ug/l   | 92       |
| 31) Cyclohexane               | 5.544          | 56   | 2061                | 1.319   | ug/l # | 46       |
| 36) 1,1-Dichloropropene       | 5.544          | 75   | 7129                | 5.546   | ug/l # | 51       |
| 38) Carbon Tetrachloride      | 5.544          | 117  | 7944                | 6.087   | ug/l # | 16       |
| 39) Methylcyclohexane         | 7.117          | 83   | 1913                | 1.242   | ug/l # | 1        |
| 42) 1,2-Dichloroethane        | 6.098          | 62   | 358                 | 0.245   | ug/l # | 75       |
| 44) Trichloroethene           | 7.117          | 130  | 129756              | 136.137 | ug/l   | 95       |
| 49) 1,4-Dioxane               | 7.684          | 88   | 247                 | 9.529   | ug/l # | 63       |
| 60) Dibromochloromethane      | 9.269          | 129  | 3537                | 3.450   | ug/l # | 11       |
| 64) Tetrachloroethene         | 9.275          | 164  | 3620                | 4.475   | ug/l   | 94       |
| 76) 1,2,3-Trichloropropane    | 11.079         | 75   | 31784               | 28.353  | ug/l # | 36       |
| 81) trans-1,4-Dichloro-2-b... | 11.079         | 75   | 31784               | 97.030  | ug/l # | 6        |
| 95) Naphthalene               | 13.786         | 128  | 861                 | 0.248   | ug/l # | 79       |
| 96) 1,2,3-Trichlorobenzene    | 13.969         | 180  | 195                 | 0.205   | ug/l # | 89       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032525\  
Data File : VX045452.D  
Acq On : 25 Mar 2025 19:51  
Operator : JC/MD  
Sample : Q1608-11  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318

Quant Time: Mar 27 03:17:47 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X022825W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 28 06:45:16 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.544 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


Tgt Ion:168 Resp: 72182  
Ion Ratio Lower Upper  
168 100  
99 65.2 48.2 72.4



#2  
Dichlorodifluoromethane  
Concen: 0.265 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion: 85 Resp: 241  
Ion Ratio Lower Upper  
85 100  
87 22.1 17.4 52.3









#5  
 Bromomethane  
 Concen: 0.295 ug/l  
 RT: 1.593 min Scan# 81  
 Delta R.T. 0.000 min  
 Lab File: VX045452.D  
 Acq: 25 Mar 2025 19:51




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE132D1-20250318


Tgt Ion: 94 Resp: 128  
 Ion Ratio Lower Upper  
 94 100  
 96 0.0 75.0 112.4#



#9  
 1,1,2-Trichlorotrifluoroethane  
 Concen: 4.924 ug/l  
 RT: 2.319 min Scan# 202  
 Delta R.T. 0.006 min  
 Lab File: VX045452.D  
 Acq: 25 Mar 2025 19:51

Tgt Ion: 101 Resp: 4131  
 Ion Ratio Lower Upper  
 101 100  
 85 46.1 36.2 54.4  
 151 70.3 56.4 84.6





#11

Tert butyl alcohol

Concen: 2.368 ug/l

RT: 2.934 min Scan# 30

Delta R.T. -0.049 min

Lab File: VX045452.D

Acq: 25 Mar 2025 19:51

Instrument :

MSVOA\_X

ClientSampleId :





RE132D1-20250318

Tgt Ion: 59 Resp: 515

Ion Ratio Lower Upper

59 100

57 0.0 7.8 11.8#



#12

1,1-Dichloroethene

Concen: 0.735 ug/l

RT: 2.307 min Scan# 200

Delta R.T. 0.000 min

Lab File: VX045452.D

Acq: 25 Mar 2025 19:51




Tgt Ion: 96 Resp: 652


Ion Ratio Lower Upper

96 100

61 179.0 134.6 202.0

98 55.1 51.0 76.6









#16  
Acetone  
Concen: 1.738 ug/l  
RT: 2.386 min Scan# 213  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


Tgt Ion: 43 Resp: 926  
Ion Ratio Lower Upper  
43 100  
58 33.3 24.2 36.4



#19  
Methyl tert-butyl Ether  
Concen: 0.757 ug/l  
RT: 3.117 min Scan# 333  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion: 73 Resp: 2254  
Ion Ratio Lower Upper  
73 100  
57 22.5 18.5 27.7









#22  
Diisopropyl ether  
Concen: 0.468 ug/l  
RT: 3.764 min Scan# 41  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 45    | Resp: | 1502  |
| Ion      | Ratio | Lower | Upper |
| 45       | 100   |       |       |
| 43       | 46.1  | 54.9  | 82.3# |
| 87       | 36.6  | 21.0  | 31.4# |
| 59       | 0.0   | 8.7   | 13.1# |



#23  
Vinyl Acetate  
Concen: 0.372 ug/l  
RT: 3.757 min Scan# 438  
Delta R.T. 0.037 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 43    | Resp: | 997   |
| Ion      | Ratio | Lower | Upper |
| 43       | 100   |       |       |
| 86       | 0.0   | 8.1   | 12.1# |









#24  
 1,1-Dichloroethane  
 Concen: 0.518 ug/l  
 RT: 3.605 min Scan# 41  
 Delta R.T. 0.000 min  
 Lab File: VX045452.D  
 Acq: 25 Mar 2025 19:51




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE132D1-20250318


Tgt Ion: 63 Resp: 932  
 Ion Ratio Lower Upper  
 63 100  
 98 0.0 3.4 10.2#  
 100 0.0 2.1 6.5#



#27  
 cis-1,2-Dichloroethene  
 Concen: 2.150 ug/l  
 RT: 4.501 min Scan# 560  
 Delta R.T. 0.018 min  
 Lab File: VX045452.D  
 Acq: 25 Mar 2025 19:51

Tgt Ion: 96 Resp: 2326  
 Ion Ratio Lower Upper  
 96 100  
 61 122.9 0.0 283.2  
 98 60.4 0.0 128.0









#28  
Bromochloromethane  
Concen: 0.281 ug/l  
RT: 5.093 min Scan# 61  
Delta R.T. 0.195 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


Tgt Ion: 49 Resp: 241  
Ion Ratio Lower Upper  
49 100  
129 0.0 0.0 3.4  
130 0.0 56.1 84.1#



#30  
Chloroform  
Concen: 2.131 ug/l  
RT: 5.093 min Scan# 657  
Delta R.T. 0.006 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion: 83 Resp: 3811  
Ion Ratio Lower Upper  
83 100  
85 58.3 51.4 77.2









#31  
Cyclohexane  
Concen: 1.319 ug/l  
RT: 5.544 min Scan# 718  
Delta R.T. 0.079 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


Tgt Ion: 56 Resp: 2061  
Ion Ratio Lower Upper  
56 100  
69 138.6 23.4 35.2#  
84 84.0 69.4 104.2



#33  
1,2-Dichloroethane-d4  
Concen: 52.824 ug/l  
RT: 5.952 min Scan# 798  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion: 65 Resp: 60650  
Ion Ratio Lower Upper  
65 100  
67 49.6 0.0 106.2









#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.757 min Scan# 91  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


Tgt Ion:114 Resp: 140177  
Ion Ratio Lower Upper  
114 100  
63 23.0 0.0 41.8  
88 16.2 0.0 32.8



#35  
Dibromofluoromethane  
Concen: 48.389 ug/l  
RT: 5.379 min Scan# 704  
Delta R.T. -0.006 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion:113 Resp: 45356  
Ion Ratio Lower Upper  
113 100  
111 102.2 81.8 122.6  
192 16.4 14.3 21.5





#36

1,1-Dichloropropene

Concen: 5.546 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.146 min

Lab File: VX045452.D

Acq: 25 Mar 2025 19:51

Instrument :

MSVOA\_X

ClientSampleId :

RE132D1-20250318





Tgt Ion: 75 Resp: 7129

Ion Ratio Lower Upper

75 100

110 9.3 16.9 50.6#

77 0.0 24.5 36.7#



#38

Carbon Tetrachloride

Concen: 6.087 ug/l

RT: 5.544 min Scan# 731

Delta R.T. -0.128 min

Lab File: VX045452.D

Acq: 25 Mar 2025 19:51




Tgt Ion: 117 Resp: 7944


Ion Ratio Lower Upper

117 100

119 5.5 76.7 115.1#


121 0.0 25.5 38.3#








#39  
Methylcyclohexane  
Concen: 1.242 ug/l  
RT: 7.117 min Scan# 91  
Delta R.T. -0.256 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318







Tgt Ion: 83 Resp: 1913  
Ion Ratio Lower Upper  
83 100  
55 0.0 63.0 94.4#  
98 157.8 39.7 59.5#



#42  
1,2-Dichloroethane  
Concen: 0.245 ug/l  
RT: 6.098 min Scan# 822  
Delta R.T. 0.018 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion: 62 Resp: 358  
Ion Ratio Lower Upper  
62 100  
98 0.0 0.0 18.2









#44  
Trichloroethene  
Concen: 136.137 ug/l  
RT: 7.117 min Scan# 91  
Delta R.T. -0.006 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


Tgt Ion:130 Resp: 129756  
Ion Ratio Lower Upper  
130 100  
95 107.8 0.0 205.0



#49  
1,4-Dioxane  
Concen: 9.529 ug/l  
RT: 7.684 min Scan# 1082  
Delta R.T. 0.024 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion: 88 Resp: 247  
Ion Ratio Lower Upper  
88 100  
43 8.9 28.7 43.1#  
58 43.3 55.8 83.8#









#50  
Toluene-d8  
Concen: 46.495 ug/l  
RT: 8.647 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


Tgt Ion: 98 Resp: 157998  
Ion Ratio Lower Upper  
98 100  
100 66.3 52.0 78.0



#60  
Dibromochloromethane  
Concen: 3.450 ug/l  
RT: 9.269 min Scan# 1342  
Delta R.T. -0.250 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion: 129 Resp: 3537  
Ion Ratio Lower Upper  
129 100  
127 0.0 38.5 115.5#









#62  
4-Bromofluorobenzene  
Concen: 48.281 ug/l  
RT: 11.079 min Scan# 1  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


Tgt Ion: 95 Resp: 54366  
Ion Ratio Lower Upper  
95 100  
174 70.4 0.0 148.2  
176 67.9 0.0 141.4



#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.049 min Scan# 1470  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion: 117 Resp: 126304  
Ion Ratio Lower Upper  
117 100  
82 61.3 46.3 69.5  
119 33.3 25.7 38.5









#64  
Tetrachloroethene  
Concen: 4.475 ug/l  
RT: 9.275 min Scan# 11  
Delta R.T. 0.006 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


|     |         |       |       |
|-----|---------|-------|-------|
| Tgt | Ion:164 | Resp: | 3620  |
| Ion | Ratio   | Lower | Upper |
| 164 | 100     |       |       |
| 166 | 133.2   | 103.6 | 155.4 |
| 129 | 95.2    | 82.7  | 124.1 |
| 131 | 92.2    | 80.5  | 120.7 |



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.018 min Scan# 1793  
Delta R.T. 0.000 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

|             |       |       |       |
|-------------|-------|-------|-------|
| Tgt Ion:152 | Resp: | 48163 |       |
| Ion         | Ratio | Lower | Upper |
| 152         | 100   |       |       |
| 115         | 62.6  | 44.2  | 132.6 |
| 150         | 156.3 | 0.0   | 349.0 |









#76  
1,2,3-Trichloropropane  
Concen: 28.353 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. -0.159 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51




Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D1-20250318


Tgt Ion: 75 Resp: 31784  
Ion Ratio Lower Upper  
75 100  
77 0.9 20.7 62.1#



#81  
trans-1,4-Dichloro-2-butene  
Concen: 97.030 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.061 min  
Lab File: VX045452.D  
Acq: 25 Mar 2025 19:51

Tgt Ion: 75 Resp: 31784  
Ion Ratio Lower Upper  
75 100  
53 0.0 88.5 132.7#  
89 0.0 36.2 54.4#





#95

Naphthalene

Concen: 0.248 ug/l

RT: 13.786 min Scan# 2081

Delta R.T. 0.012 min

Lab File: VX045452.D

Acq: 25 Mar 2025 19:51

Instrument :

MSVOA\_X

ClientSampleId :

RE132D1-20250318





Tgt Ion:128 Resp: 861

Ion Ratio Lower Upper

128 100

127 7.1 10.3 15.5#

129 0.0 8.7 13.1#



#96

1,2,3-Trichlorobenzene

Concen: 0.205 ug/l

RT: 13.969 min Scan# 2113

Delta R.T. 0.012 min

Lab File: VX045452.D

Acq: 25 Mar 2025 19:51

Tgt Ion:180 Resp: 195

Ion Ratio Lower Upper

180 100

182 93.8 47.5 142.6

145 12.8 17.0 50.9#

