

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
 Data File : VX045435.D
 Acq On : 25 Mar 2025 13:12
 Operator : JC/MD
 Sample : Q1608-10DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP08-20250318DL

Quant Time: Mar 26 01:36:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133107	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49927	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58279	54.727	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.460%	
35) Dibromofluoromethane	5.385	113	43017	48.331	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.660%	
50) Toluene-d8	8.647	98	154249	47.803	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.600%	
62) 4-Bromofluorobenzene	11.079	95	53940	50.447	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 100.900%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	2331	2.996	ug/l	97
44) Trichloroethene	7.117	130	33559	37.080	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
Data File : VX045435.D
Acq On : 25 Mar 2025 13:12
Operator : JC/MD
Sample : Q1608-10DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP08-20250318DL

Quant Time: Mar 26 01:36:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

