

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
 Data File : VX045438.D
 Acq On : 25 Mar 2025 14:26
 Operator : JC/MD
 Sample : Q1608-19DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D7-20250318DL

Quant Time: Mar 26 01:38:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	72406	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	142524	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62108	53.926	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.860%
35) Dibromofluoromethane	5.385	113	47495	49.836	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.680%
50) Toluene-d8	8.647	98	163588	47.348	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.700%
62) 4-Bromofluorobenzene	11.079	95	58454	51.056	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.120%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.123	130	64366	66.419	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
Data File : VX045438.D
Acq On : 25 Mar 2025 14:26
Operator : JC/MD
Sample : Q1608-19DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D7-20250318DL

Quant Time: Mar 26 01:38:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

