

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
 Data File : VX045439.D
 Acq On : 25 Mar 2025 14:50
 Operator : JC/MD
 Sample : Q1608-21DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D2-20250319DL

Quant Time: Mar 26 01:39:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	70652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	139172	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	131487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54151	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61817	55.006	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.020%
35) Dibromofluoromethane	5.379	113	45901	49.324	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.640%
50) Toluene-d8	8.647	98	163272	48.394	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.780%
62) 4-Bromofluorobenzene	11.079	95	58937	52.718	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.440%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.313	101	694	0.845	ug/l	97
12) 1,1-Dichloroethene	2.307	96	227	0.261	ug/l #	78
44) Trichloroethene	7.117	130	34101	36.036	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
Data File : VX045439.D
Acq On : 25 Mar 2025 14:50
Operator : JC/MD
Sample : Q1608-21DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D2-20250319DL

Quant Time: Mar 26 01:39:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

