

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
 Data File : VX045444.D
 Acq On : 25 Mar 2025 16:46
 Operator : JC/MD
 Sample : Q1629-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB03-20250321

Quant Time: Mar 26 01:41:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	71073	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	141090	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49151	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62649	55.416	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.840%
35) Dibromofluoromethane	5.379	113	46667	49.465	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.940%
50) Toluene-d8	8.647	98	163786	47.887	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.780%
62) 4-Bromofluorobenzene	11.079	95	56202	49.588	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.180%
Target Compounds						
16) Acetone	2.380	43	2084	3.973	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
Data File : VX045444.D
Acq On : 25 Mar 2025 16:46
Operator : JC/MD
Sample : Q1629-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB03-20250321

Quant Time: Mar 26 01:41:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

