

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
 Data File : VX045450.D
 Acq On : 25 Mar 2025 19:05
 Operator : JC/MD
 Sample : Q1629-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT150S1-20250320

Quant Time: Mar 26 01:43:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	142479	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62550	54.636	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.280%	
35) Dibromofluoromethane	5.379	113	47321	49.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.340%	
50) Toluene-d8	8.647	98	163441	47.320	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 94.640%	
62) 4-Bromofluorobenzene	11.079	95	58211	50.860	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.720%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1314	2.474	ug/l	94
44) Trichloroethene	7.123	130	11230	11.592	ug/l	92
64) Tetrachloroethene	9.275	164	885	1.057	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
Data File : VX045450.D
Acq On : 25 Mar 2025 19:05
Operator : JC/MD
Sample : Q1629-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT150S1-20250320

Quant Time: Mar 26 01:43:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

