

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
 Data File : VX045431.D
 Acq On : 25 Mar 2025 11:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 26 01:33:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.629	0.798	-26.9#	103	0.00
3 P	Chloromethane	0.770	0.804	-4.4	86	0.00
4 C	Vinyl Chloride	0.764	0.720	5.8#	76	0.00
5 T	Bromomethane	0.300	0.346	-15.3	95	0.00
6 T	Chloroethane	0.352	0.394	-11.9	85	0.00
7 T	Trichlorofluoromethane	1.011	1.093	-8.1	84	0.00
8 T	Diethyl Ether	0.396	0.356	10.1	74	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.581	0.699	-20.3	95	0.00
10 T	Methyl Iodide	0.726	0.823	-13.4	86	0.00
11 T	Tert butyl alcohol	0.151	0.133	11.9	80	0.00
12 CM	1,1-Dichloroethene	0.614	0.651	-6.0#	84	0.00
13 T	Acrolein	0.174	0.187	-7.5	87	0.00
14 T	Allyl chloride	1.094	1.222	-11.7	89	0.00
15 T	Acrylonitrile	0.404	0.404	0.0	80	0.00
16 T	Acetone	0.369	0.394	-6.8	90	0.00
17 T	Carbon Disulfide	1.636	1.643	-0.4	80	0.00
18 T	Methyl Acetate	0.888	0.856	3.6	80	0.00
19 T	Methyl tert-butyl Ether	2.061	2.180	-5.8	84	0.00
20 T	Methylene Chloride	0.731	0.729	0.3	84	0.00
21 T	trans-1,2-Dichloroethene	0.607	0.641	-5.6	82	0.00
22 T	Diisopropyl ether	2.224	2.316	-4.1	82	0.00
23 T	Vinyl Acetate	1.856	2.127	-14.6	87	0.00
24 P	1,1-Dichloroethane	1.247	1.311	-5.1	85	0.00
25 T	2-Butanone	0.555	0.596	-7.4	83	0.00
26 T	2,2-Dichloropropane	0.585	0.969	-65.6#	131	0.00
27 T	cis-1,2-Dichloroethene	0.749	0.773	-3.2	82	0.00
28 T	Bromochloromethane	0.594	0.584	1.7	78	0.00
29 T	Tetrahydrofuran	0.373	0.374	-0.3	81	0.00
30 C	Chloroform	1.239	1.317	-6.3#	85	0.00
31 T	Cyclohexane	1.082	1.188	-9.8	87	0.00
32 T	1,1,1-Trichloroethane	1.000	1.132	-13.2	89	0.00
33 S	1,2-Dichloroethane-d4	0.795	0.876	-10.2	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	78	0.00
35 S	Dibromofluoromethane	0.334	0.368	-10.2	87	0.00
36 T	1,1-Dichloropropene	0.459	0.515	-12.2	85	0.00
37 T	Ethyl Acetate	0.591	0.640	-8.3	82	0.00
38 T	Carbon Tetrachloride	0.465	0.558	-20.0	93	0.00
39 T	Methylcyclohexane	0.549	0.674	-22.8	90	0.00
40 TM	Benzene	1.448	1.547	-6.8	80	0.00
41 T	Methacrylonitrile	0.319	0.356	-11.6	81	0.00
42 TM	1,2-Dichloroethane	0.521	0.624	-19.8	92	0.00
43 T	Isopropyl Acetate	0.874	1.008	-15.3	84	0.00
44 TM	Trichloroethene	0.340	0.365	-7.4	83	0.00
45 C	1,2-Dichloropropane	0.372	0.390	-4.8#	82	0.00
46 T	Dibromomethane	0.263	0.302	-14.8	87	0.00
47 T	Bromodichloromethane	0.516	0.591	-14.5	88	0.00
48 T	Methyl methacrylate	0.438	0.523	-19.4	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
 Data File : VX045431.D
 Acq On : 25 Mar 2025 11:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 26 01:33:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	81	0.00
50 S	Toluene-d8	1.212	1.260	-4.0	81	0.00
51 T	4-Methyl-2-Pentanone	0.587	0.681	-16.0	87	0.00
52 CM	Toluene	0.849	0.944	-11.2#	82	0.00
53 T	t-1,3-Dichloropropene	0.431	0.547	-26.9#	91	0.00
54 T	cis-1,3-Dichloropropene	0.503	0.610	-21.3	89	0.00
55 T	1,1,2-Trichloroethane	0.350	0.370	-5.7	81	0.00
56 T	Ethyl methacrylate	0.532	0.624	-17.3	85	0.00
57 T	1,3-Dichloropropane	0.605	0.651	-7.6	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.309	-19.3	85	0.00
59 T	2-Hexanone	0.425	0.519	-22.1	90	0.00
60 T	Dibromochloromethane	0.366	0.425	-16.1	86	0.00
61 T	1,2-Dibromoethane	0.349	0.390	-11.7	85	0.00
62 S	4-Bromofluorobenzene	0.402	0.481	-19.7	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.320	0.353	-10.3	87	0.00
65 PM	Chlorobenzene	1.058	1.130	-6.8	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.352	0.388	-10.2	87	0.00
67 C	Ethyl Benzene	1.843	2.068	-12.2#	85	0.00
68 T	m/p-Xylenes	0.675	0.760	-12.6	84	0.00
69 T	o-Xylene	0.681	0.753	-10.6	85	0.00
70 T	Styrene	1.101	1.269	-15.3	86	0.00
71 P	Bromoform	0.266	0.306	-15.0	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.903	4.076	-4.4	87	0.00
74 T	N-amyl acetate	1.831	2.109	-15.2	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.432	1.422	0.7	89	0.00
76 T	1,2,3-Trichloropropane	1.164	1.204	-3.4	93	0.00
77 T	Bromobenzene	0.921	0.914	0.8	86	0.00
78 T	n-propylbenzene	4.409	4.877	-10.6	90	0.00
79 T	2-Chlorotoluene	2.808	2.956	-5.3	91	0.00
80 T	1,3,5-Trimethylbenzene	3.156	3.383	-7.2	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.340	0.390	-14.7	102	0.00
82 T	4-Chlorotoluene	3.060	3.387	-10.7	93	0.00
83 T	tert-Butylbenzene	3.242	3.440	-6.1	90	0.00
84 T	1,2,4-Trimethylbenzene	3.175	3.497	-10.1	92	0.00
85 T	sec-Butylbenzene	3.884	4.330	-11.5	91	0.00
86 T	p-Isopropyltoluene	3.139	3.628	-15.6	94	0.00
87 T	1,3-Dichlorobenzene	1.643	1.755	-6.8	93	0.00
88 T	1,4-Dichlorobenzene	1.662	1.731	-4.2	91	0.00
89 T	n-Butylbenzene	2.703	3.340	-23.6	98	0.00
90 T	Hexachloroethane	0.568	0.624	-9.9	91	0.00
91 T	1,2-Dichlorobenzene	1.648	1.733	-5.2	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.323	-15.8	99	0.00
93 T	1,2,4-Trichlorobenzene	0.938	1.047	-11.6	92	0.00
94 T	Hexachlorobutadiene	0.385	0.429	-11.4	97	0.00
95 T	Naphthalene	3.605	3.876	-7.5	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032525\
Data File : VX045431.D
Acq On : 25 Mar 2025 11:31
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 26 01:33:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.990	1.048	-5.9	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6