

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032619\
 Data File : VX008447.D
 Acq On : 26 Mar 2019 11:14
 Operator : JC/SP
 Sample : VX0326WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0326WBL01

Quant Time: Mar 27 02:02:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	216766	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	364033	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	307610	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126009	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	154322	46.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.92%	
35) Dibromofluoromethane	5.49	113	113773	47.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.86%	
50) Toluene-d8	8.71	98	453837	47.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.64%	
62) 4-Bromofluorobenzene	11.13	95	151107	42.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.92%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032619\
Data File : VX008447.D
Acq On : 26 Mar 2019 11:14
Operator : JC/SP
Sample : VX0326WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0326WBL01

Quant Time: Mar 27 02:02:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

