

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032619\
 Data File : VX008459.D
 Acq On : 26 Mar 2019 17:09
 Operator : JC/SP
 Sample : I.BLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I.BLK

Quant Time: Mar 27 03:09:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	218724	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	366773	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	312730	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131819	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	153970	46.43	ug/l	0.00
Spiked Amount			Recovery	=	92.86%	
35) Dibromofluoromethane	5.49	113	112210	46.92	ug/l	0.00
Spiked Amount			Recovery	=	93.84%	
50) Toluene-d8	8.71	98	457745	47.87	ug/l	0.00
Spiked Amount			Recovery	=	95.74%	
62) 4-Bromofluorobenzene	11.13	95	155301	43.83	ug/l	0.00
Spiked Amount			Recovery	=	87.66%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032619\
Data File : VX008459.D
Acq On : 26 Mar 2019 17:09
Operator : JC/SP
Sample : I.BLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I.BLK

Quant Time: Mar 27 03:09:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

