

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032619\
 Data File : VX008466.D
 Acq On : 26 Mar 2019 19:54
 Operator : JC/SP
 Sample : K2089-01DL 200X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-106DL

Quant Time: Mar 27 03:30:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	241560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	406291	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	351034	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151653	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	167729	45.80	ug/l	0.00
Spiked Amount			Recovery	=	91.60%	
35) Dibromofluoromethane	5.49	113	124069	46.83	ug/l	0.00
Spiked Amount			Recovery	=	93.66%	
50) Toluene-d8	8.71	98	511869	48.32	ug/l	0.00
Spiked Amount			Recovery	=	96.64%	
62) 4-Bromofluorobenzene	11.13	95	176712	45.02	ug/l	0.00
Spiked Amount			Recovery	=	90.04%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	6558	3.426	ug/l	100
25) 2-Butanone	4.68	43	7085	2.309	ug/l	95
31) Cyclohexane	5.57	56	4494	0.766	ug/l	95
39) Methylcyclohexane	7.46	83	2138	0.413	ug/l	95
40) Benzene	6.14	78	123863	9.551	ug/l	100
52) Toluene	8.78	92	789812	101.853	ug/l	99
67) Ethyl Benzene	10.25	91	76131	5.284	ug/l	98
68) m/p-Xylenes	10.35	106	147818	28.202	ug/l	100
69) o-Xylene	10.69	106	62653	12.318	ug/l	100
73) Isopropylbenzene	11.02	105	5343	0.433	ug/l	99
78) n-propylbenzene	11.36	91	13547	0.937	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	20549	1.970	ug/l	97
84) 1,2,4-Trimethylbenzene	11.80	105	79153	7.503	ug/l	99
95) Naphthalene	13.83	128	25100	2.030	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032619\
Data File : VX008466.D
Acq On : 26 Mar 2019 19:54
Operator : JC/SP
Sample : K2089-01DL 200X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-106DL

Quant Time: Mar 27 03:30:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration

