

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032619\
 Data File : VX008449.D
 Acq On : 26 Mar 2019 12:11
 Operator : JC/SP
 Sample : K2082-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 27 02:07:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 07:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	230545	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	393335	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	338517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	140334	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	166470	47.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.26%	
35) Dibromofluoromethane	5.49	113	121794	47.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.98%	
50) Toluene-d8	8.71	98	493882	48.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.32%	
62) 4-Bromofluorobenzene	11.13	95	166076	43.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.40%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032619\
Data File : VX008449.D
Acq On : 26 Mar 2019 12:11
Operator : JC/SP
Sample : K2082-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 27 02:07:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 07:57:20 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. 0.00 min

Lab File: VX008449.D

Acq: 26 Mar 2019 12:11

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-WATER-REF-5

Tot Ion:168 Resp: 230545

Ion Ratio Lower Upper

168 100

99 59.4 46.0 69.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 47.629 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX008449.D


Acq: 26 Mar 2019 12:11

Tgt Ion: 65 Resp: 166470

Ion Ratio Lower Upper

65 100

67 54.0 0.0 109.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX008449.D
 Acq: 26 Mar 2019 12:11

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Tot Ion:114 Resp: 393335
 Ion Ratio Lower Upper
 114 100
 63 22.7 0.0 46.4
 88 16.9 0.0 33.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 47.488 ug/l
 RT: 5.49 min Scan# 712
 Delta R.T. 0.00 min
 Lab File: VX008449.D
 Acq: 26 Mar 2019 12:11

Tgt Ion:113 Resp: 121794
 Ion Ratio Lower Upper
 113 100
 111 102.1 83.0 124.4
 192 19.3 15.6 23.4



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.159 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX008449.D
Acq: 26 Mar 2019 12:11




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tot Ion: 98 Resp: 493882
Ion Ratio Lower Upper
98 100
100 63.8 51.1 76.7



#62
4-Bromofluorobenzene
Concen: 43.702 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.00 min
Lab File: VX008449.D
Acq: 26 Mar 2019 12:11

Tgt Ion: 95 Resp: 166076
Ion Ratio Lower Upper
95 100
174 74.2 0.0 147.2
176 72.0 0.0 142.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX008449.D
Acq: 26 Mar 2019 12:11

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Tot Ion:117 Resp: 338517
Ion Ratio Lower Upper
117 100
82 62.1 49.4 74.2
119 32.3 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX008449.D
Acq: 26 Mar 2019 12:11

Tgt Ion:152 Resp: 140334
Ion Ratio Lower Upper
152 100
115 61.3 43.5 130.5
150 156.9 0.0 348.8

