

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032620\
 Data File : VX015358.D
 Acq On : 26 Mar 2020 15:07
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 27 08:13:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 26 02:31:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.98	128	197368	30.00	ug/l	-0.01
28) 1,4-Difluorobenzene	6.83	114	1107866	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	977904	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.03	65	462062	29.96	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery =	99.87%		
60) 4-Bromofluorobenzene	11.12	95	508034	30.36	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery =	101.20%		
63) Toluene-d8	8.70	98	1277024	30.74	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery =	102.47%		

Target Compounds

					Ovalue
83) 1,3-Dichlorobenzene	12.01	146	10738	0.401 ug/l	93
89) 1,2,4-Trichlorobenzene	13.63	180	22188	1.054 ug/l	92
90) Hexachlorobutadiene	13.77	225	13449	1.213 ug/l	98
91) Naphthalene	13.82	128	53555	0.879 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032620\
Data File : VX015358.D
Acq On : 26 Mar 2020 15:07
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 27 08:13:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X032520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 26 02:31:53 2020
Response via : Initial Calibration

