

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021535.D
 Acq On : 26 Mar 2021 16:14
 Operator : JC/MD
 Sample : M1458-07 2.5PPB LOD
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 27 04:52:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032621W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Mar 27 04:47:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.982	128	13742	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.829	114	76717	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.094	117	68432	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.029	65	38368	29.90	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.67%
60) 4-Bromofluorobenzene	11.118	95	34685	28.82	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	96.07%
63) Toluene-d8	8.694	98	94488	30.74	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.47%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.187	85	2164	1.69	ug/l	91
3) Chloromethane	1.311	50	3030	2.60	ug/l	94
4) Vinyl Chloride	1.393	62	2214	2.13	ug/l	99
5) Bromomethane	1.623	94	1201	2.88	ug/l	90
6) Chloroethane	1.705	64	1370	2.21	ug/l #	86
7) Trichlorofluoromethane	1.911	101	3411	2.19	ug/l	92
8) Diethyl Ether	2.164	74	1117	2.18	ug/l	43
9) 1,1,2-Trichlorotrifluo...	2.364	101	1836	2.13	ug/l	96
10) 1,1-Dichloroethene	2.346	96	1760	2.15	ug/l	93
11) Methyl Iodide	2.487	142	1400	2.05	ug/l	94
12) Methyl Acetate	2.752	43	2953	1.76	ug/l #	85
13) Acrolein	2.276	56	2022	12.82	ug/l	96
14) Acrylonitrile	3.117	53	5463	10.47	ug/l	88
15) Acetone	2.423	58	1684	12.19	ug/l	94
16) Carbon Disulfide	2.546	76	4966	2.17	ug/l	97
17) Allyl chloride	2.705	41	3679	2.32	ug/l	84
18) Methylene Chloride	2.834	84	2533	2.62	ug/l #	90
19) trans-1,2-Dichloroethene	3.135	96	2002	2.26	ug/l #	90
20) Diisopropyl ether	3.823	45	7079	2.23	ug/l #	92
21) 1,1-Dichloroethane	3.664	63	3977	2.31	ug/l #	90
22) cis-1,2-Dichloroethene	4.564	96	2340	2.29	ug/l #	81
23) tert-Butyl Alcohol	3.017	59	2615	11.81	ug/l #	17
24) Methyl tert-Butyl Ether	3.170	73	6823	2.28	ug/l #	84
25) Chloroform	5.170	83	4048	2.21	ug/l	93
26) Cyclohexane	5.541	56	3259	2.19	ug/l #	44
29) 1,1-Dichloropropene	5.758	75	2899	2.24	ug/l #	56
30) 2-Butanone	4.641	43	8794	11.63	ug/l	92
31) 2,2-Dichloropropane	4.547	77	3194	2.21	ug/l #	80
32) 1,1,1-Trichloroethane	5.464	97	3634	2.27	ug/l #	75
33) Carbon Tetrachloride	5.735	117	1181	0.83	ug/l #	75
34) Benzene	6.111	78	8157	2.21	ug/l	93
35) Methacrylonitrile	5.000	41	1130	1.36	ug/l	87
36) 1,2-Dichloroethane	6.158	62	3849	2.43	ug/l #	88
37) Trichloroethene	7.182	130	2252	2.35	ug/l	89
38) Methylcyclohexane	7.441	83	3055	2.13	ug/l	90
39) 1,2-Dichloropropane	7.494	63	2413	2.45	ug/l	90
40) Dibromomethane	7.629	93	1737	2.41	ug/l	85
41) Bromodichloromethane	7.876	83	3271	2.21	ug/l	99
42) Vinyl Acetate	3.782	43	29325	11.29	ug/l #	92


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021535.D
 Acq On : 26 Mar 2021 16:14
 Operator : JC/MD
 Sample : M1458-07 2.5PPB LOD
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 27 04:52:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032621W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Mar 27 04:47:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	4.799	43	3394	2.36	ug/l #	71
44) Isopropyl Acetate	6.406	43	5433	2.25	ug/l #	75
45) 1,4-Dioxane	7.718	88	1006	41.93	ug/l #	81
46) Methyl methacrylate	7.741	41	2717	2.05	ug/l	88
47) n-amyl Acetate	10.877	43	4140	1.99	ug/l	96
48) t-1,3-Dichloropropene	9.024	75	3069	2.06	ug/l	99
49) cis-1,3-Dichloropropene	8.412	75	3288	2.06	ug/l #	84
50) 1,1,2-Trichloroethane	9.194	97	2210	2.29	ug/l	98
51) Ethyl methacrylate	9.159	69	2987	2.11	ug/l #	70
52) 1,3-Dichloropropane	9.353	76	3794	2.30	ug/l	98
53) Dibromochloromethane	9.565	129	2299	2.09	ug/l	97
54) 1,2-Dibromoethane	9.653	107	2331	2.25	ug/l	96
55) 2-Chloroethyl vinyl ether	8.288	63	9387	11.95	ug/l	96
56) Bromoform	10.841	173	1628	2.09	ug/l	99
58) 4-Methyl-2-Pentanone	8.618	43	16622	11.51	ug/l #	88
59) 2-Hexanone	9.471	43	12130	11.18	ug/l #	86
61) Tetrachloroethene	9.318	164	1967	2.37	ug/l	96
62) Toluene	8.765	91	8866	2.30	ug/l	97
64) Chlorobenzene	10.118	112	5675	2.41	ug/l	96
65) 1,1,1,2-Tetrachloroethane	10.200	131	2089	2.25	ug/l	97
66) Ethyl Benzene	10.235	91	9180	2.13	ug/l	93
67) m/p-Xylenes	10.341	106	6632	4.18	ug/l	91
68) o-Xylene	10.683	106	3330	2.17	ug/l	95
69) Styrene	10.694	104	5492	2.15	ug/l	95
70) Isopropylbenzene	11.000	105	8788	2.10	ug/l	98
71) 1,1,2,2-Tetrachloroethane	11.247	83	3374	2.36	ug/l	98
72) 1,2,3-Trichloropropane	11.277	75	3778	2.76	ug/l	91
73) Bromobenzene	11.236	156	2198	2.18	ug/l	91
74) n-propylbenzene	11.342	91	10326	2.11	ug/l	96
75) 2-Chlorotoluene	11.400	91	6531	2.19	ug/l	98
76) 1,3,5-Trimethylbenzene	11.489	105	7447	2.09	ug/l	100
77) t-1,4-Dichloro-2-butene	11.053	75	796	1.82	ug/l	60
78) 4-Chlorotoluene	11.494	91	7453	2.14	ug/l	96
79) tert-butylbenzene	11.753	119	7312	2.16	ug/l	94
80) 1,2,4-Trimethylbenzene	11.789	105	7423	2.09	ug/l	96
81) sec-Butylbenzene	11.930	105	7944	2.02	ug/l	99
82) p-Isopropyltoluene	12.047	119	7014	1.95	ug/l	94
83) 1,3-Dichlorobenzene	12.006	146	4002	2.22	ug/l	97
84) 1,4-Dichlorobenzene	12.083	146	3861	2.14	ug/l	95
85) n-Butylbenzene	12.371	91	6018	1.88	ug/l	96
86) Hexachloroethane	12.577	117	1489	2.36	ug/l	85
87) 1,2-Dichlorobenzene	12.371	146	3819	2.15	ug/l	96
88) 1,2-Dibromo-3-Chloropr...	12.983	75	729	2.18	ug/l	87
89) 1,2,4-Trichlorobenzene	13.630	180	2341	2.09	ug/l	99
90) Hexachlorobutadiene	13.765	225	1264	2.67	ug/l	89
91) Naphthalene	13.818	128	7158	1.92	ug/l	99
92) 1,2,3-Trichlorobenzene	14.001	180	2333	2.09	ug/l	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 30.00 ug/l
RT: 4.982 min Scan# 662
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:128 Resp: 13742
Ion Ratio Lower Upper
128 100
49 185.2 0.0 367.5
130 132.3 0.0 318.8



#2
Dichlorodifluoromethane
Concen: 1.69 ug/l
RT: 1.187 min Scan# 17
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 85 Resp: 2164
Ion Ratio Lower Upper
85 100
87 28.3 16.8 50.4









#3
Chloromethane
Concen: 2.60 ug/l
RT: 1.311 min Scan# 38
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 50 Resp: 3030
Ion Ratio Lower Upper
50 100
52 30.2 26.9 40.3



#4
Vinyl Chloride
Concen: 2.13 ug/l
RT: 1.393 min Scan# 52
Delta R.T. 0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 62 Resp: 2214
Ion Ratio Lower Upper
62 100
64 32.3 25.4 38.2









#5
Bromomethane
Concen: 2.88 ug/l
RT: 1.623 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 1201
Ion Ratio Lower Upper
94 100
96 86.1 76.6 114.8



#6
Chloroethane
Concen: 2.21 ug/l
RT: 1.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 64 Resp: 1370
Ion Ratio Lower Upper
64 100
66 26.0 27.4 41.2#









#7
Trichlorofluoromethane
Concen: 2.19 ug/l
RT: 1.911 min Scan# 140
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 3411
Ion Ratio Lower Upper
101 100
103 58.0 51.4 77.2



#8
Diethyl Ether
Concen: 2.18 ug/l
RT: 2.164 min Scan# 183
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 74 Resp: 1117
Ion Ratio Lower Upper
74 100
45 138.9 17.3 155.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 2.13 ug/l
RT: 2.364 min Scan# 217
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 1836
Ion Ratio Lower Upper
101 100
85 48.3 0.0 88.0
151 73.1 0.0 151.8



#10
1,1-Dichloroethene
Concen: 2.15 ug/l
RT: 2.346 min Scan# 214
Delta R.T. -0.012 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14


Tgt Ion: 96 Resp: 1760
Ion Ratio Lower Upper
96 100
61 165.8 126.1 189.1
98 58.5 51.3 76.9








#11
Methyl Iodide
Concen: 2.05 ug/l
RT: 2.487 min Scan# 238
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:142 Resp: 1400
Ion Ratio Lower Upper
142 100
127 45.1 21.4 64.2
141 18.0 6.5 19.4



#12
Methyl Acetate
Concen: 1.76 ug/l
RT: 2.752 min Scan# 283
Delta R.T. 0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 43 Resp: 2953
Ion Ratio Lower Upper
43 100
74 20.7 22.8 34.2#









#13
Acrolein
Concen: 12.82 ug/l
RT: 2.276 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 2022
Ion Ratio Lower Upper
56 100
55 73.4 56.3 84.5



#14
Acrylonitrile
Concen: 10.47 ug/l
RT: 3.117 min Scan# 345
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 53 Resp: 5463
Ion Ratio Lower Upper
53 100
52 93.2 41.1 123.3
51 41.6 17.6 52.9









#15
Acetone
Concen: 12.19 ug/l
RT: 2.423 min Scan# 227
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 58 Resp: 1684
Ion Ratio Lower Upper
58 100
43 302.9 232.6 348.8



#16
Carbon Disulfide
Concen: 2.17 ug/l
RT: 2.546 min Scan# 248
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14


Tgt Ion: 76 Resp: 4966
Ion Ratio Lower Upper
76 100
78 8.2 7.4 11.0








#17
 Allyl chloride
 Concen: 2.32 ug/l
 RT: 2.705 min Scan# 275
 Delta R.T. 0.000 min
 Lab File: VX021535.D
 Acq: 26 Mar 2021 16:14

Instrument :
 MSVOA_X
 ClientSampled :







Tgt Ion:	41	Resp:	3679
Ion	Ratio	Lower	Upper
41	100		
39	85.2	34.8	104.4
76	36.4	21.3	63.9



#18
 Methylene Chloride
 Concen: 2.62 ug/l
 RT: 2.834 min Scan# 297
 Delta R.T. -0.001 min
 Lab File: VX021535.D
 Acq: 26 Mar 2021 16:14


Tgt Ion:	84	Resp:	2533
Ion	Ratio	Lower	Upper
84	100		
49	122.8	90.6	136.0
51	44.2	26.3	39.5#
86	59.1	51.3	76.9








#19
trans-1,2-Dichloroethene
Concen: 2.26 ug/l
RT: 3.135 min Scan# 348
Delta R.T. -0.005 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 2002
Ion Ratio Lower Upper
96 100
61 135.8 103.0 154.4
98 50.6 51.8 77.6#




#20
Diisopropyl ether
Concen: 2.23 ug/l
RT: 3.823 min Scan# 465
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14



Tgt Ion: 45 Resp: 7079
Ion Ratio Lower Upper
45 100
43 76.2 56.0 84.0
87 27.4 25.4 38.0
59 9.7 10.1 15.1#





#21

1,1-Dichloroethane

Concen: 2.31 ug/l

RT: 3.664 min Scan# 438





Delta R.T. -0.006 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14

Tgt Ion:	63	Resp:	3977
Ion Ratio	Lower	Upper	
63	100		
98	5.5	3.9	11.7
100	0.0	2.5	7.4#

Instrument :
MSVOA_X
ClientSampleId :



#22

cis-1,2-Dichloroethene

Concen: 2.29 ug/l




RT: 4.564 min Scan# 591


Delta R.T. -0.006 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14

Tgt Ion:	96	Resp:	2340
Ion Ratio	Lower	Upper	
96	100		
61	158.9	114.1	171.1
98	40.5	51.8	77.8#









#23
tert-Butyl Alcohol
Concen: 11.81 ug/l
RT: 3.017 min Scan# 328
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 59 Resp: 2615
Ion Ratio Lower Upper
59 100
52 194.8 0.1 0.1#



#24
Methyl tert-Butyl Ether
Concen: 2.28 ug/l
RT: 3.170 min Scan# 354
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 73 Resp: 6823
Ion Ratio Lower Upper
73 100
43 25.8 14.9 22.3#





#25

Chloroform

Concen: 2.21 ug/l

RT: 5.170 min Scan# 694

Delta R.T. -0.006 min

Lab File: VX021535.D





Acq: 26 Mar 2021 16:14

Tgt Ion: 83 Resp: 4048

Ion Ratio Lower Upper

83 100

85 69.2 51.2 76.8



#26

Cyclohexane

Concen: 2.19 ug/l

RT: 5.541 min Scan# 757

Delta R.T. -0.012 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14




Tgt Ion: 56 Resp: 3259


Ion Ratio Lower Upper

56 100

89 0.0 0.0 0.0

84 38.6 74.1 111.1#





#27

1,2-Dichloroethane-d4

Concen: 29.90 ug/l

RT: 6.029 min Scan# 840

Delta R.T. 0.000 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14

Tgt Ion: 65 Resp: 38368

Ion Ratio Lower Upper





65 100

67 50.4 41.9 62.9

Instrument :

MSVOA_X

ClientSampled :



#28

1,4-Difluorobenzene

Concen: 30.00 ug/l

RT: 6.829 min Scan# 976

Delta R.T. -0.006 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14




Tgt Ion: 114 Resp: 76717


Ion Ratio Lower Upper

114 100

63 22.8 16.2 24.2

88 16.5 13.0 19.6





#29

1,1-Dichloropropene

Concen: 2.24 ug/l

RT: 5.758 min Scan# 794

Delta R.T. -0.012 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14





Tgt Ion: 75 Resp: 2899

Ion Ratio Lower Upper

75 100

110 16.3 0.0 72.4

77 0.0 0.0 62.4



#30

2-Butanone

Concen: 11.63 ug/l

RT: 4.641 min Scan# 604

Delta R.T. 0.006 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14




Tgt Ion: 43 Resp: 8794


Ion Ratio Lower Upper

43 100

57 8.1 7.2 10.8

72 24.6 23.7 35.5









#31
2,2-Dichloropropane
Concen: 2.21 ug/l
RT: 4.547 min Scan# 588
Delta R.T. -0.005 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 77 Resp: 3194
Ion Ratio Lower Upper
77 100
97 13.5 18.6 28.0#



#32
1,1,1-Trichloroethane
Concen: 2.27 ug/l
RT: 5.464 min Scan# 744
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 97 Resp: 3634
Ion Ratio Lower Upper
97 100
99 34.4 51.7 77.5#
61 46.8 34.3 51.5









#33
Carbon Tetrachloride
Concen: 0.83 ug/l
RT: 5.735 min Scan# 790
Delta R.T. -0.018 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 1181
Ion Ratio Lower Upper
117 100
119 66.6 76.1 114.1#
121 24.3 24.1 36.1



#34
Benzene
Concen: 2.21 ug/l
RT: 6.111 min Scan# 854
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 78 Resp: 8157
Ion Ratio Lower Upper
78 100
77 20.4 19.0 28.4
51 13.0 12.5 18.7





#35

Methacrylonitrile

Concen: 1.36 ug/l

RT: 5.000 min Scan# 665

Delta R.T. -0.012 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14

Tgt Ion: 41 Resp: 1130





Ion Ratio Lower Upper

41 100

39 56.1 26.8 80.3

67 60.4 39.1 117.2

52 22.4 14.0 42.0



#36

1,2-Dichloroethane

Concen: 2.43 ug/l

RT: 6.158 min Scan# 862

Delta R.T. -0.006 min

Lab File: VX021535.D




Acq: 26 Mar 2021 16:14


Tgt Ion: 62 Resp: 3849

Ion Ratio Lower Upper

62 100

98 5.6 8.1 12.1#





#37

Trichloroethene

Concen: 2.35 ug/l

RT: 7.182 min Scan# 1036

Delta R.T. -0.006 min

Lab File: VX021535.D





Acq: 26 Mar 2021 16:14

Tgt Ion:130 Resp: 2252

Ion Ratio Lower Upper

130 100

95 89.1 80.2 120.4



#38

Methylcyclohexane

Concen: 2.13 ug/l

RT: 7.441 min Scan# 1080

Delta R.T. -0.000 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14




Tgt Ion: 83 Resp: 3055


Ion Ratio Lower Upper

83 100

55 83.5 37.4 112.1

98 42.9 24.7 74.1





#39

1,2-Dichloropropane

Concen: 2.45 ug/l

RT: 7.494 min Scan# 1089

Delta R.T. 0.006 min

Lab File: VX021535.D





Acq: 26 Mar 2021 16:14

Tgt Ion: 63 Resp: 2413

Ion Ratio Lower Upper

63 100

65 26.5 25.6 38.4



#40

Dibromomethane

Concen: 2.41 ug/l

RT: 7.629 min Scan# 1112

Delta R.T. -0.012 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14




Tgt Ion: 93 Resp: 1737


Ion Ratio Lower Upper

93 100

95 76.2 0.0 167.4

174 84.5 0.0 211.8









#41
Bromodichloromethane
Concen: 2.21 ug/l
RT: 7.876 min Scan# 1154
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 3271
Ion Ratio Lower Upper
83 100
85 64.7 51.4 77.0
127 7.6 7.1 10.7



#42
Vinyl Acetate
Concen: 11.29 ug/l
RT: 3.782 min Scan# 458
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 43 Resp: 29325
Ion Ratio Lower Upper
43 100
86 8.3 9.0 13.4#





#43

Ethyl Acetate

Concen: 2.36 ug/l

RT: 4.799 min Scan# 631

Delta R.T. 0.006 min

Lab File: VX021535.D





Acq: 26 Mar 2021 16:14

Tgt Ion: 43 Resp: 3394

Ion Ratio Lower Upper

43 100

45 3.3 12.2 18.4#



#44

Isopropyl Acetate

Concen: 2.25 ug/l

RT: 6.406 min Scan# 904

Delta R.T. -0.006 min

Lab File: VX021535.D




Acq: 26 Mar 2021 16:14


Tgt Ion: 43 Resp: 5433

Ion Ratio Lower Upper

43 100

61 10.4 18.1 27.1#









#45
1,4-Dioxane
Concen: 41.93 ug/l
RT: 7.718 min Scan# 1127
Delta R.T. 0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 88 Resp: 1006
Ion Ratio Lower Upper
88 100
43 42.4 22.2 33.2#
58 79.6 53.9 80.9



#46
Methyl methacrylate
Concen: 2.05 ug/l
RT: 7.741 min Scan# 1131
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 41 Resp: 2717
Ion Ratio Lower Upper
41 100
69 76.1 46.9 140.6
39 54.0 28.0 83.9









#47
n-amy1 Acetate
Concen: 1.99 ug/l
RT: 10.877 min Scan# 1664
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 4140
Ion Ratio Lower Upper
43 100
55 25.3 21.8 32.6



#48
t-1,3-Dichloropropene
Concen: 2.06 ug/l
RT: 9.024 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 75 Resp: 3069
Ion Ratio Lower Upper
75 100
77 30.7 25.0 37.6









#49

cis-1,3-Dichloropropene
Concen: 2.06 ug/l
RT: 8.412 min Scan# 1245
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 75 Resp: 3288
Ion Ratio Lower Upper
75 100
77 33.0 24.4 36.6
39 60.0 35.4 53.0#




#50

1,1,2-Trichloroethane
Concen: 2.29 ug/l
RT: 9.194 min Scan# 1378
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 97 Resp: 2210
Ion Ratio Lower Upper
97 100
83 82.6 68.5 102.7
85 55.1 43.8 65.6
99 59.8 49.4 74.2









#51
Ethyl methacrylate
Concen: 2.11 ug/l
RT: 9.159 min Scan# 1372
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 69 Resp: 2987
Ion Ratio Lower Upper
69 100
41 76.8 29.1 87.4
39 52.3 15.4 46.1#



#52
1,3-Dichloropropane
Concen: 2.30 ug/l
RT: 9.353 min Scan# 1405
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 76 Resp: 3794
Ion Ratio Lower Upper
76 100
78 31.8 0.0 66.0









#53
Dibromochloromethane
Concen: 2.09 ug/l
RT: 9.565 min Scan# 1441
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:129 Resp: 2299
Ion Ratio Lower Upper
129 100
127 74.6 38.4 115.2



#54
1,2-Dibromoethane
Concen: 2.25 ug/l
RT: 9.653 min Scan# 1456
Delta R.T. 0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion:107 Resp: 2331
Ion Ratio Lower Upper
107 100
109 95.5 73.6 110.4









#55
2-Chloroethyl vinyl ether
Concen: 11.95 ug/l
RT: 8.288 min Scan# 1224
Delta R.T. 0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:	63	Resp:	9387
Ion	Ratio	Lower	Upper
63	100		
65	32.2	25.9	38.9
106	26.0	23.9	35.9



#56
Bromoform
Concen: 2.09 ug/l
RT: 10.841 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion:	173	Resp:	1628
Ion	Ratio	Lower	Upper
173	100		
175	47.7	38.1	57.1
254	9.5	8.2	12.4









#57
Chlorobenzene-d5
Concen: 30.00 ug/l
RT: 10.094 min Scan# 1531
Delta R.T. 0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 68432
Ion Ratio Lower Upper
117 100
82 60.2 45.0 67.6
119 32.1 25.6 38.4



#58
4-Methyl-2-Pentanone
Concen: 11.51 ug/l
RT: 8.618 min Scan# 1280
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 43 Resp: 16622
Ion Ratio Lower Upper
43 100
58 35.7 35.2 52.8
85 16.6 16.8 25.2#









#59
2-Hexanone
Concen: 11.18 ug/l
RT: 9.471 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 12130
Ion Ratio Lower Upper
43 100
58 49.0 49.0 73.6#
57 18.4 17.1 25.7



#60
4-Bromofluorobenzene
Concen: 28.82 ug/l
RT: 11.118 min Scan# 1705
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 95 Resp: 34685
Ion Ratio Lower Upper
95 100
174 73.4 63.0 94.4
176 69.9 59.9 89.9









#61
Tetrachloroethene
Concen: 2.37 ug/l
RT: 9.318 min Scan# 1399
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	164	Resp:	1967
Ion	Ratio	Lower	Upper
164	100		
166	122.7	102.8	154.2
129	97.3	75.8	113.8
131	93.9	77.8	116.6



#62
Toluene
Concen: 2.30 ug/l
RT: 8.765 min Scan# 1305
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14


Tgt Ion:	91	Resp:	8866
Ion	Ratio	Lower	Upper
91	100		
92	57.0	47.2	70.8








#63
Toluene-d8
Concen: 30.74 ug/l
RT: 8.694 min Scan# 1293
Delta R.T. 0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Instrument :
MSVOA_X
ClientSampled :







Tgt Ion: 98 Resp: 94488
Ion Ratio Lower Upper
98 100
100 64.1 53.0 79.4



#64
Chlorobenzene
Concen: 2.41 ug/l
RT: 10.118 min Scan# 1535
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 112 Resp: 5675
Ion Ratio Lower Upper
112 100
114 29.9 25.7 38.5









#65
1,1,1,2-Tetrachloroethane
Concen: 2.25 ug/l
RT: 10.200 min Scan# 1549
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:131 Resp: 2089
Ion Ratio Lower Upper
131 100
133 98.5 0.0 189.6
119 66.8 0.0 129.0



#66
Ethyl Benzene
Concen: 2.13 ug/l
RT: 10.235 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 91 Resp: 9180
Ion Ratio Lower Upper
91 100
106 27.2 25.0 37.4









#67
m/p-Xylenes
Concen: 4.18 ug/l
RT: 10.341 min Scan# 1573
Delta R.T. 0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 6632
Ion Ratio Lower Upper
106 100
91 218.7 164.0 246.0



#68
o-Xylene
Concen: 2.17 ug/l
RT: 10.683 min Scan# 1631
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion:106 Resp: 3330
Ion Ratio Lower Upper
106 100
91 221.1 106.5 319.6





#69

Styrene

Concen: 2.15 ug/l

RT: 10.694 min Scan# 1633





Delta R.T. -0.000 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14

Tgt Ion:	104	Resp:	5492
Ion	Ratio	Lower	Upper
104	100		
78	56.3	39.2	58.8
103	53.5	43.0	64.4

Instrument :
MSVOA_X
ClientSampleId :



#70

Isopropylbenzene

Concen: 2.10 ug/l




RT: 11.000 min Scan# 1685


Delta R.T. -0.000 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14

Tgt Ion:	105	Resp:	8788
Ion	Ratio	Lower	Upper
105	100		
120	25.2	18.4	34.2









#71
1,1,2,2-Tetrachloroethane
Concen: 2.36 ug/l
RT: 11.247 min Scan# 1727
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 3374
Ion Ratio Lower Upper
83 100
131 10.5 5.1 15.3
85 65.3 32.0 96.0



#72
1,2,3-Trichloropropane
Concen: 2.76 ug/l
RT: 11.277 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 75 Resp: 3778
Ion Ratio Lower Upper
75 100
77 32.0 0.0 74.6





#73

Bromobenzene

Concen: 2.18 ug/l

RT: 11.236 min Scan# 1725





Delta R.T. -0.000 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14

Tgt Ion:	156	Resp:	2198
Ion	Ratio	Lower	Upper
156	100		
77	179.1	0.0	324.8
158	99.9	0.0	192.0

Instrument :
MSVOA_X
ClientSampled :



#74

n-propylbenzene

Concen: 2.11 ug/l




RT: 11.342 min Scan# 1743


Delta R.T. -0.000 min

Lab File: VX021535.D

Acq: 26 Mar 2021 16:14

Tgt Ion:	91	Resp:	10326
Ion	Ratio	Lower	Upper
91	100		
120	20.9	0.0	45.8









#75
2-Chlorotoluene
Concen: 2.19 ug/l
RT: 11.400 min Scan# 1753
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 6531
Ion Ratio Lower Upper
91 100
126 32.4 0.0 66.6



#76
1,3,5-Trimethylbenzene
Concen: 2.09 ug/l
RT: 11.489 min Scan# 1768
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 105 Resp: 7447
Ion Ratio Lower Upper
105 100
120 48.2 0.0 96.2









#77
t-1,4-Dichloro-2-butene
Concen: 1.82 ug/l
RT: 11.053 min Scan# 1694
Delta R.T. -0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 796
Ion Ratio Lower Upper
75 100
53 139.2 48.4 145.0
89 65.9 22.7 68.0



#78
4-Chlorotoluene
Concen: 2.14 ug/l
RT: 11.494 min Scan# 1769
Delta R.T. -0.001 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 91 Resp: 7453
Ion Ratio Lower Upper
91 100
126 26.9 0.0 57.8









#79
tert-butylbenzene
Concen: 2.16 ug/l
RT: 11.753 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:119 Resp: 7312
Ion Ratio Lower Upper
119 100
91 64.1 0.0 118.0
134 22.5 0.0 48.0



#80
1,2,4-Trimethylbenzene
Concen: 2.09 ug/l
RT: 11.789 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion:105 Resp: 7423
Ion Ratio Lower Upper
105 100
120 42.1 0.0 89.8









#81
 sec-Butylbenzene
 Concen: 2.02 ug/l
 RT: 11.930 min Scan# 1843
 Delta R.T. -0.000 min
 Lab File: VX021535.D
 Acq: 26 Mar 2021 16:14




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion:105 Resp: 7944
 Ion Ratio Lower Upper
 105 100
 134 20.2 0.0 39.6



#82
 p-Isopropyltoluene
 Concen: 1.95 ug/l
 RT: 12.047 min Scan# 1863
 Delta R.T. -0.000 min
 Lab File: VX021535.D
 Acq: 26 Mar 2021 16:14

Tgt Ion:119 Resp: 7014
 Ion Ratio Lower Upper
 119 100
 134 24.3 0.0 52.8
 91 27.6 0.0 47.4









#83
1,3-Dichlorobenzene
Concen: 2.22 ug/l
RT: 12.006 min Scan# 1856
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 4002
Ion Ratio Lower Upper
146 100
111 42.7 20.3 60.9
148 66.2 31.9 95.9



#84
1,4-Dichlorobenzene
Concen: 2.14 ug/l
RT: 12.083 min Scan# 1869
Delta R.T. 0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion:146 Resp: 3861
Ion Ratio Lower Upper
146 100
111 44.2 19.1 57.3
148 65.8 31.9 95.5









#85
n-Butylbenzene
Concen: 1.88 ug/l
RT: 12.371 min Scan# 1918
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	91	Resp:	6018
Ion	Ratio	Lower	Upper
91	100		
92	51.7	0.0	109.0
134	23.4	0.0	51.8



#86
Hexachloroethane
Concen: 2.36 ug/l
RT: 12.577 min Scan# 1953
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion:	117	Resp:	1489
Ion	Ratio	Lower	Upper
117	100		
201	68.0	40.5	121.4









#87
1,2-Dichlorobenzene
Concen: 2.15 ug/l
RT: 12.371 min Scan# 1918
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 3819
Ion Ratio Lower Upper
146 100
111 43.9 20.8 62.5
148 68.6 32.4 97.0



#88
1,2-Dibromo-3-Chloropropane
Concen: 2.18 ug/l
RT: 12.983 min Scan# 2022
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion: 75 Resp: 729
Ion Ratio Lower Upper
75 100
155 76.0 68.9 103.3
157 97.9 91.3 136.9









#89
1,2,4-Trichlorobenzene
Concen: 2.09 ug/l
RT: 13.630 min Scan# 2132
Delta R.T. 0.006 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:180 Resp: 2341
Ion Ratio Lower Upper
180 100
182 98.3 77.9 116.9
145 28.9 24.3 36.5



#90
Hexachlorobutadiene
Concen: 2.67 ug/l
RT: 13.765 min Scan# 2155
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion:225 Resp: 1264
Ion Ratio Lower Upper
225 100
223 56.3 0.0 129.4
227 56.3 0.0 130.4









#91
Naphthalene
Concen: 1.92 ug/l
RT: 13.818 min Scan# 2164
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:	128	Resp:	7158
Ion	Ratio	Lower	Upper
128	100		
127	12.1	10.2	15.4
129	11.2	8.8	13.2



#92
1,2,3-Trichlorobenzene
Concen: 2.09 ug/l
RT: 14.001 min Scan# 2195
Delta R.T. -0.000 min
Lab File: VX021535.D
Acq: 26 Mar 2021 16:14

Tgt Ion:	180	Resp:	2333
Ion	Ratio	Lower	Upper
180	100		
182	97.2	0.0	192.4
145	33.6	0.0	66.4

