

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021544.D
 Acq On : 26 Mar 2021 21:38
 Operator : JC/MD
 Sample : M1817-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2021-03-24-TB-01

Quant Time: Mar 27 04:32:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	60995	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.829	114	112110	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	114361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	49020	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52459	51.69	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.38%	
35) Dibromofluoromethane	5.458	113	40803	51.24	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.48%	
50) Toluene-d8	8.694	98	142307	50.15	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.30%	
62) 4-Bromofluorobenzene	11.118	95	57952	52.53	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.06%	
Target Compounds						
16) Acetone	2.423	43	19155	49.92	ug/l	Qvalue 89
25) 2-Butanone	4.646	43	2452	4.22	ug/l	95
37) Ethyl Acetate	4.805	43	1007	0.79	ug/l	# 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
Data File : VX021544.D
Acq On : 26 Mar 2021 21:38
Operator : JC/MD
Sample : M1817-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2021-03-24-TB-01

Quant Time: Mar 27 04:32:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

