

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021548.D
 Acq On : 26 Mar 2021 23:12
 Operator : JC/MD
 Sample : M1787-12DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-122D1-20210319DL

Quant Time: Mar 27 04:34:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	65512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	119161	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	123447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	55549	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	55103	50.55	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.10%
35) Dibromofluoromethane	5.464	113	41670	49.23	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.46%
50) Toluene-d8	8.694	98	151423	50.20	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.40%
62) 4-Bromofluorobenzene	11.118	95	63422	54.08	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.16%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.188	130	42996	49.14	ug/l	98
64) Tetrachloroethene	9.318	164	219	0.26	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
Data File : VX021548.D
Acq On : 26 Mar 2021 23:12
Operator : JC/MD
Sample : M1787-12DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-122D1-20210319DL

Quant Time: Mar 27 04:34:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

