

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021549.D
 Acq On : 26 Mar 2021 23:35
 Operator : JC/MD
 Sample : M1787-17 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20210319

Quant Time: Mar 27 04:34:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	67792	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	122709	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	126498	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	58362	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	57157	50.67	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.34%
35) Dibromofluoromethane	5.464	113	43526	49.94	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.88%
50) Toluene-d8	8.694	98	154905	49.87	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.74%
62) 4-Bromofluorobenzene	11.118	95	65903	54.57	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.14%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.358	101	558	0.80	ug/l	95
44) Trichloroethene	7.188	130	1721	1.91	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
Data File : VX021549.D
Acq On : 26 Mar 2021 23:35
Operator : JC/MD
Sample : M1787-17 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20210319

Quant Time: Mar 27 04:34:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

