

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021550.D
 Acq On : 26 Mar 2021 23:59
 Operator : JC/MD
 Sample : M1787-11DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-122D2-20210319DL

Quant Time: Mar 27 04:34:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	67151	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	122926	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	124199	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	56419	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	57633	51.58	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.16%		
35) Dibromofluoromethane	5.464	113	43503	49.82	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.64%		
50) Toluene-d8	8.694	98	154275	49.58	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.16%		
62) 4-Bromofluorobenzene	11.118	95	62137	51.36	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 102.72%		
Target Compounds						Qvalue
44) Trichloroethene	7.188	130	49207	54.52	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
Data File : VX021550.D
Acq On : 26 Mar 2021 23:59
Operator : JC/MD
Sample : M1787-11DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-122D2-20210319DL

Quant Time: Mar 27 04:34:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

