

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021552.D
 Acq On : 27 Mar 2021 00:46
 Operator : JC/MD
 Sample : M1787-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D-20210318

Quant Time: Mar 27 04:35:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	65845	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	118726	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	116370	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	51748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	55698	50.84	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.68%
35) Dibromofluoromethane	5.470	113	42181	50.02	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.04%
50) Toluene-d8	8.694	98	150725	50.15	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.30%
62) 4-Bromofluorobenzene	11.118	95	59538	50.96	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.92%
Target Compounds						
16) Acetone	2.423	43	1682	4.06	ug/l	96
44) Trichloroethene	7.188	130	2500	2.87	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
Data File : VX021552.D
Acq On : 27 Mar 2021 00:46
Operator : JC/MD
Sample : M1787-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D-20210318

Quant Time: Mar 27 04:35:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

