

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021553.D
 Acq On : 27 Mar 2021 01:09
 Operator : JC/MD
 Sample : M1787-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-126D1-20210319

Quant Time: Mar 27 04:35:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	64084	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	114897	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	118689	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	53408	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	54312	50.94	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.88%	
35) Dibromofluoromethane	5.464	113	41260	50.55	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.10%	
50) Toluene-d8	8.694	98	145706	50.10	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.20%	
62) 4-Bromofluorobenzene	11.118	95	59505	52.63	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.26%	
Target Compounds						
						Qvalue
44) Trichloroethene	7.188	130	33493	39.70	ug/l	95
49) 1,4-Dioxane	7.718	88	311	15.41	ug/l #	77
64) Tetrachloroethene	9.312	164	867	1.09	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
Data File : VX021553.D
Acq On : 27 Mar 2021 01:09
Operator : JC/MD
Sample : M1787-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-126D1-20210319

Quant Time: Mar 27 04:35:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

