

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021557.D
 Acq On : 27 Mar 2021 02:43
 Operator : JC/MD
 Sample : M1787-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW203D2-20210320

Quant Time: Mar 27 04:36:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	59110	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	109570	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	108634	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	43009	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	51449	52.31	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.62%
35) Dibromofluoromethane	5.470	113	39700	51.01	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.02%
50) Toluene-d8	8.694	98	137429	49.55	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.10%
62) 4-Bromofluorobenzene	11.118	95	52134	48.35	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.70%
Target Compounds						
					Qvalue	
16) Acetone	2.423	43	1361	3.66	ug/l #	84
24) 1,1-Dichloroethane	3.670	63	4330	3.42	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
Data File : VX021557.D
Acq On : 27 Mar 2021 02:43
Operator : JC/MD
Sample : M1787-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW203D2-20210320

Quant Time: Mar 27 04:36:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

