

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021559.D
 Acq On : 27 Mar 2021 03:30
 Operator : JC/MD
 Sample : M1787-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW202D-20210320

Quant Time: Mar 27 04:36:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	58494	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	108491	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110349	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	49832	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	51002	52.40	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.80%
35) Dibromofluoromethane	5.464	113	38564	50.04	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.08%
50) Toluene-d8	8.694	98	135741	49.43	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.86%
62) 4-Bromofluorobenzene	11.118	95	56408	52.83	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.66%
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	1.917	101	2054	1.92	ug/l	89
16) Acetone	2.423	43	1103	3.00	ug/l	98
19) Methyl tert-butyl Ether	3.170	73	1249	0.57	ug/l	97
22) Diisopropyl ether	3.834	45	1206	0.52	ug/l #	74
24) 1,1-Dichloroethane	3.670	63	5460	4.36	ug/l	97
27) cis-1,2-Dichloroethene	4.564	96	5520	7.34	ug/l	98
42) 1,2-Dichloroethane	6.170	62	2063	1.58	ug/l	91
44) Trichloroethene	7.188	130	27837	34.95	ug/l	89
64) Tetrachloroethene	9.318	164	6795	9.19	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
Data File : VX021559.D
Acq On : 27 Mar 2021 03:30
Operator : JC/MD
Sample : M1787-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW202D-20210320

Quant Time: Mar 27 04:36:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

