



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032621\
 Data File : VX021524.D
 Acq On : 26 Mar 2021 10:43
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032621W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Sat Mar 27 04:47:38 2021



AutoFind: Scans 1704, 1705, 1706; Background Corrected with Scan 1699

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.2	3817	PASS
75	95	30	60	55.8	9604	PASS
95	95	100	100	100.0	17224	PASS
96	95	5	9	6.5	1127	PASS
173	174	0.00	2	1.5	177	PASS
174	95	50	100	69.9	12032	PASS
175	174	5	9	7.1	860	PASS
176	174	95	101	95.2	11458	PASS
177	176	5	9	6.8	779	PASS