

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032622\
 Data File : VX027719.D
 Acq On : 26 Mar 2022 11:55
 Operator : JC/MD
 Sample : VX0326WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0326WBL01

Quant Time: Mar 28 05:12:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	62827	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	113005	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40184	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54302	52.697	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.400%	
35) Dibromofluoromethane	5.391	113	38668	52.107	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.220%	
50) Toluene-d8	8.653	98	141250	49.239	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.480%	
62) 4-Bromofluorobenzene	11.085	95	51622	44.348	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 88.700%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032622\
Data File : VX027719.D
Acq On : 26 Mar 2022 11:55
Operator : JC/MD
Sample : VX0326WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0326WBL01

Quant Time: Mar 28 05:12:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 16:26:59 2022
Response via : Initial Calibration

