

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032622\
 Data File : VX027724.D
 Acq On : 26 Mar 2022 14:04
 Operator : JC/MD
 Sample : N2057-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 UST-7-BOTTOM

Quant Time: Mar 28 05:15:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63093	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	111888	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39399	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46318	44.760	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.520%
35) Dibromofluoromethane	5.385	113	36922	50.251	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.500%
50) Toluene-d8	8.653	98	135860	47.833	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.660%
62) 4-Bromofluorobenzene	11.085	95	51152	44.383	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.760%
Target Compounds						
16) Acetone	2.386	43	1282	2.929	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032622\
Data File : VX027724.D
Acq On : 26 Mar 2022 14:04
Operator : JC/MD
Sample : N2057-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
UST-7-BOTTOM

Quant Time: Mar 28 05:15:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 16:26:59 2022
Response via : Initial Calibration

