

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032622\
 Data File : VX027730.D
 Acq On : 26 Mar 2022 16:26
 Operator : JC/MD
 Sample : N2085-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-3-23-2022

Quant Time: Mar 28 09:03:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63863	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	112998	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99905	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41478	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52797	50.406	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.820%			
35) Dibromofluoromethane	5.391	113	38764	52.239	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 104.480%			
50) Toluene-d8	8.653	98	139756	48.721	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 97.440%			
62) 4-Bromofluorobenzene	11.085	95	55000	47.253	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 94.500%			
Target Compounds						
16) Acetone	2.380	43	13043	29.443	ug/l	93
25) 2-Butanone	4.580	43	2642	3.714	ug/l #	70
40) Benzene	6.050	78	999	0.309	ug/l #	51
51) 4-Methyl-2-Pentanone	8.586	43	685	0.484	ug/l #	61
52) Toluene	8.720	92	2211	1.092	ug/l	86
67) Ethyl Benzene	10.195	91	3963	1.011	ug/l #	87
68) m/p-Xylenes	10.299	106	4692	3.302	ug/l	90
69) o-Xylene	10.640	106	2654	1.859	ug/l	86
73) Isopropylbenzene	10.969	105	803	0.236	ug/l	99
78) n-propylbenzene	11.305	91	1064	0.263	ug/l	77
80) 1,3,5-Trimethylbenzene	11.451	105	1812	0.619	ug/l	90
84) 1,2,4-Trimethylbenzene	11.756	105	5617	1.890	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032622\
Data File : VX027730.D
Acq On : 26 Mar 2022 16:26
Operator : JC/MD
Sample : N2085-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-3-23-2022

Quant Time: Mar 28 09:03:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 16:26:59 2022
Response via : Initial Calibration

