

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX032622\
 Data File : VX027717.D
 Acq On : 26 Mar 2022 10:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 28 05:11:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	50.000	61.289	-22.6	95	0.00
3 P	Chloromethane	50.000	49.692	0.6	87	0.00
4 C	Vinyl Chloride	50.000	53.661	-7.3#	90	0.00
5 T	Bromomethane	50.000	51.500	-3.0	99	0.00
6 T	Chloroethane	50.000	49.135	1.7	93	0.00
7 T	Trichlorofluoromethane	50.000	54.758	-9.5	97	0.00
8 T	Diethyl Ether	50.000	56.341	-12.7	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	55.045	-10.1	101	0.00
10 T	Methyl Iodide	50.000	58.179	-16.4	102	0.00
11 T	Tert butyl alcohol	250.000	278.511	-11.4	91	0.00
12 CM	1,1-Dichloroethene	50.000	52.415	-4.8#	93	0.00
13 T	Acrolein	250.000	272.585	-9.0	100	0.00
14 T	Allyl chloride	50.000	54.797	-9.6	91	0.00
15 T	Acrylonitrile	250.000	248.374	0.7	82	0.00
16 T	Acetone	250.000	275.010	-10.0	105	0.00
17 T	Carbon Disulfide	50.000	50.251	-0.5	91	0.00
18 T	Methyl Acetate	50.000	55.628	-11.3	92	0.00
19 T	Methyl tert-butyl Ether	50.000	52.692	-5.4	88	0.00
20 T	Methylene Chloride	50.000	55.149	-10.3	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.053	-0.1	92	0.00
22 T	Diisopropyl ether	50.000	55.266	-10.5	97	0.00
23 T	Vinyl Acetate	250.000	281.911	-12.8	95	0.00
24 P	1,1-Dichloroethane	50.000	54.777	-9.6	94	0.00
25 T	2-Butanone	250.000	215.786	13.7	78	0.00
26 T	2,2-Dichloropropane	50.000	51.519	-3.0	94	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.151	-4.3	91	0.00
28 T	Bromochloromethane	50.000	51.493	-3.0	89	0.00
29 T	Tetrahydrofuran	250.000	274.189	-9.7	90	0.00
30 C	Chloroform	50.000	51.169	-2.3#	93	0.00
31 T	Cyclohexane	50.000	53.701	-7.4	98	0.00
32 T	1,1,1-Trichloroethane	50.000	54.223	-8.4	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	51.293	-2.6	99	-0.01
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	84	0.00
35 S	Dibromofluoromethane	50.000	50.650	-1.3	98	0.00
36 T	1,1-Dichloropropene	50.000	52.870	-5.7	98	0.00
37 T	Ethyl Acetate	50.000	45.211	9.6	78	0.00
38 T	Carbon Tetrachloride	50.000	54.780	-9.6	99	0.00
39 T	Methylcyclohexane	50.000	55.158	-10.3	94	0.00
40 TM	Benzene	50.000	53.525	-7.0	91	0.00
41 T	Methacrylonitrile	50.000	46.803	6.4	77	0.00
42 TM	1,2-Dichloroethane	50.000	53.023	-6.0	88	0.00
43 T	Isopropyl Acetate	50.000	55.765	-11.5	88	0.00
44 TM	Trichloroethene	50.000	50.987	-2.0	94	0.00
45 C	1,2-Dichloropropane	50.000	55.210	-10.4#	94	0.00
46 T	Dibromomethane	50.000	55.015	-10.0	93	0.00
47 T	Bromodichloromethane	50.000	55.450	-10.9	92	0.00
48 T	Methyl methacrylate	50.000	60.264	-20.5	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032622\
 Data File : VX027717.D
 Acq On : 26 Mar 2022 10:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 28 05:11:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 16:26:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1113.078	-11.3	99	0.00
50 S	Toluene-d8	50.000	46.136	7.7	88	0.00
51 T	4-Methyl-2-Pentanone	250.000	221.489	11.4	73	0.00
52 CM	Toluene	50.000	52.463	-4.9#	91	0.00
53 T	t-1,3-Dichloropropene	50.000	53.449	-6.9	96	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.964	-3.9	85	0.00
55 T	1,1,2-Trichloroethane	50.000	55.623	-11.2	96	0.00
56 T	Ethyl methacrylate	50.000	57.791	-15.6	93	0.00
57 T	1,3-Dichloropropane	50.000	57.122	-14.2	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	262.684	-5.1	85	0.00
59 T	2-Hexanone	250.000	285.010	-14.0	95	0.00
60 T	Dibromochloromethane	50.000	59.738	-19.5	99	0.00
61 T	1,2-Dibromoethane	50.000	54.785	-9.6	95	0.00
62 S	4-Bromofluorobenzene	50.000	49.220	1.6	93	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	84	0.00
64 T	Tetrachloroethene	50.000	63.582	-27.2#	101	0.00
65 PM	Chlorobenzene	50.000	52.562	-5.1	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	58.724	-17.4	102	0.00
67 C	Ethyl Benzene	50.000	54.189	-8.4#	93	0.00
68 T	m/p-Xylenes	100.000	110.381	-10.4	94	0.00
69 T	o-Xylene	50.000	54.485	-9.0	95	0.00
70 T	Styrene	50.000	57.477	-15.0	96	0.00
71 P	Bromoform	50.000	59.580	-19.2	101	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	89	0.00
73 T	Isopropylbenzene	50.000	52.981	-6.0	94	0.00
74 T	N-amyl acetate	50.000	55.933	-11.9	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	53.313	-6.6	99	0.00
76 T	1,2,3-Trichloropropane	50.000	55.558	-11.1	100	0.00
77 T	Bromobenzene	50.000	52.923	-5.8	100	0.00
78 T	n-propylbenzene	50.000	54.390	-8.8	96	0.00
79 T	2-Chlorotoluene	50.000	52.365	-4.7	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.892	-7.8	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	57.771	-15.5	100	0.00
82 T	4-Chlorotoluene	50.000	53.700	-7.4	98	0.00
83 T	tert-Butylbenzene	50.000	53.196	-6.4	95	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.628	-7.3	97	0.00
85 T	sec-Butylbenzene	50.000	54.000	-8.0	97	0.00
86 T	p-Isopropyltoluene	50.000	56.219	-12.4	98	0.00
87 T	1,3-Dichlorobenzene	50.000	54.477	-9.0	100	0.00
88 T	1,4-Dichlorobenzene	50.000	51.689	-3.4	98	0.00
89 T	n-Butylbenzene	50.000	55.943	-11.9	100	0.00
90 T	Hexachloroethane	50.000	54.517	-9.0	97	0.00
91 T	1,2-Dichlorobenzene	50.000	54.000	-8.0	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.847	-9.7	103	0.00
93 T	1,2,4-Trichlorobenzene	50.000	56.050	-12.1	92	0.00
94 T	Hexachlorobutadiene	50.000	53.403	-6.8	93	0.00
95 T	Naphthalene	50.000	55.912	-11.8	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032622\
Data File : VX027717.D
Acq On : 26 Mar 2022 10:27
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 28 05:11:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 16:26:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	55.014	-10.0	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6