



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032622\
 Data File : VX027716.D
 Acq On : 26 Mar 2022 09:26
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032522W.M
 Title : SW846 8260
 Last Update : Fri Mar 25 16:26:59 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.6	4049	PASS
75	95	30	60	57.9	10847	PASS
95	95	100	100	100.0	18745	PASS
96	95	5	9	7.1	1340	PASS
173	174	0.00	2	1.3	141	PASS
174	95	50	100	58.1	10886	PASS
175	174	5	9	7.8	850	PASS
176	174	95	101	96.3	10482	PASS
177	176	5	9	7.3	765	PASS